

LB132460 09/16/24.


INSTRUMENT ID: CV1

SFAM01.1

Linear

 μ Abs.:

32301



A= 0.0000e+000

B= 3.1470e-004

C= -1.6502e-001

Rho= 0.9999132

Accept=Accepted

slope

Y-INTERCEPT

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/2
0.00	0.000	0.022	0.022	595	0.000	595					1
0.05	0.050					442	0				1
0.20	0.200	0.158	-0.042	1026	0.0 %	1026					-21
2.50	2.500	2.570	0.070	8692	0.0 %	8692					3
5.00	5.000	4.921	-0.079	16163	0.0 %	16163					-2
7.50	7.500	7.528	0.028	24446	0.0 %	24446					0
10.0	10.000	10.000	0.000	32301	0.0 %	32301					0

LB132460
INSTRUMENT ID: CV1

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0	595	-		0 SFAM01.1	PPB	9/16/2024 21:16	Std
0.2 S01	1026	-		0.2 SFAM01.1	PPB	9/16/2024 21:19	Std
2.5 S02	8692	-		2.5 SFAM01.1	PPB	9/16/2024 21:21	Std
5 S03	16163	-		5 SFAM01.1	PPB	9/16/2024 21:23	Std
7.5 S04	24446	-		7.5 SFAM01.1	PPB	9/16/2024 21:25	Std
10 S05	32301	-		10 SFAM01.1	PPB	9/16/2024 21:34	Std
ICV001	ICV001	13393	4.0497 -	SFAM01.1	PPB	9/16/2024 21:41	SMPL
ICB001	ICB001	724	0.0628 -	SFAM01.1	PPB	9/16/2024 21:46	SMPL
CCV001	CCV001	16027	4.8786 -	SFAM01.1	PPB	9/16/2024 21:48	SMPL
CCB001	CCB001	-23	-0.1723 -	SFAM01.1	PPB	9/16/2024 21:50	SMPL
PB163428E	PBW428	731	0.065 -	SFAM01.1	PPB	9/16/2024 21:52	SMPL
P3873-01	MCOSP2	-821	-0.4234 -	SFAM01.1	PPB	9/16/2024 21:55	SMPL
P3873-02	MCOSP2D	-242	-0.2412 -	SFAM01.1	PPB	9/16/2024 21:57	SMPL
P3873-03	MCOSP2S	2788	0.7124 -	SFAM01.1	PPB	9/16/2024 21:59	SMPL
P3935-01	MBHMK7	-185	-0.2232 -	SFAM01.1	PPB	9/16/2024 22:02	SMPL
P3935-02	MBHMK7D	637	0.0354 -	SFAM01.1	PPB	9/16/2024 22:04	SMPL
P3935-03	MBHMK7S	3042	0.7923 -	SFAM01.1	PPB	9/16/2024 22:06	SMPL
PB163430E	PBW430	75	-0.1414 -	SFAM01.1	PPB	9/16/2024 22:11	SMPL
P3921-14	YE0S5	-441	-0.3038 -	SFAM01.1	PPB	9/16/2024 22:14	SMPL
P3931-17	YE0S4	226	-0.0939 -	SFAM01.1	PPB	9/16/2024 22:16	SMPL
P3969-01	EIKS7	154	-0.1166 -	SFAM01.1	PPB	9/16/2024 22:18	SMPL
P3969-02	EIKS7D	237	-0.0904 -	SFAM01.1	PPB	9/16/2024 22:21	SMPL
P3969-03	EIKS7S	2782	0.7105 -	SFAM01.1	PPB	9/16/2024 22:23	SMPL
PB163299T	LEB299	688	0.0515 -	SFAM01.1	PPB	9/16/2024 22:25	SMPL
CCV002	CCV002	15938	4.8506 -	SFAM01.1	PPB	9/16/2024 22:27	SMPL
CCB002	CCB002	680	0.049 -	SFAM01.1	PPB	9/16/2024 22:30	SMPL