

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132474

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/18/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	09/18/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	09/18/2024	08:17
4	ICV	1	Water	NA	NA	20.3	7.02	09/18/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	09/18/2024	08:22
6	P4085-05	1	Water	NA	NA	20.1	7.30	09/18/2024	08:27
7	P4096-01	1	Water	NA	NA	20.1	6.60	09/18/2024	08:33
8	P4097-03	1	Water	NA	NA	23.3	7.75	09/18/2024	08:38
9	P4097-03DUP	1	Water	NA	NA	23.4	7.76	09/18/2024	08:39
10	CCV2	1	Water	NA	NA	20.3	12.02	09/18/2024	08:40

WORKLIST(Hardcopy Internal Chain)

VB 132474

WorkList Name : PH P4097

WorkList ID : 183587

Department : Wet-Chemistry

Date : 09-18-2024 08:00:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4085-05 Q	PURGE-WATER	Water	pH	Cool 4 deg C	PSEG03	H53	09/17/2024	9040C
P4096-01 B	DRAIN-WATER-TANK-1	Water	pH	Cool 4 deg C	MAJO01	H53	09/17/2024	9040C
P4097-03 C	FM648	Water	pH	Cool 4 deg C	PSEG03	H53	09/17/2024	9040C

Date/Time 09-18-24 08:10
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 09-18-24 10:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]