

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132502

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.1	09/19/2024	09:00
2	P4030-02		1	11	100.00	>212.0	>212.0	09/19/2024	09:30
3	P4030-02DUP		1	11	100.00	>212.0	>212.0	09/19/2024	10:20
4	P4085-05		1	14	100.00	>212.0	>212.0	09/19/2024	10:50
5	P4097-03		1	12	100.00	>212.0	>212.0	09/19/2024	11:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66132502

WorkList Name : fp-9-17 WorkList ID : 183550 Department : Wet-Chemistry Date : 09-17-2024 08:30:49

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4030-02	WC-A-08-202409	Water	Flash Point	Cool 4 deg C	AECO02	H53	09/16/2024	1010B
P4085-05	PURGE-WATER	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/17/2024	1010B
P4097-03	FM648	Water	Flash Point	Cool 4 deg C	PSEG03	H53	09/17/2024	1010B

Date/Time 09/19/2024 08:50
Raw Sample Received by: RIM
Raw Sample Relinquished by: RIM

Date/Time 09/19/2024 12:16
Raw Sample Received by: Mace
Raw Sample Relinquished by: RIM