


LB132509, INSTRUMENT ID: CV-1

7470A



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.0	0.000	-0.026	-0.026	52	0.000	52	0			1.0
0.2	0.200	0.169	-0.031	828	0.0 %	828				-16
2.5	2.500	2.468	-0.032	9971	0.0 %	9971				-1
5.0	5.000	5.069	0.069	20318	0.0 %	20318				2
7.5	7.500	7.719	0.219	30858	0.0 %	30858				3
10.0	10.000	9.800	-0.200	39135	0.0 %	39135				-2

LB132509, INSTRUMENT ID:CV1

Method: 7470A		Operator: Admin		Date of Analysis: 19 Sep 2024 11:19:26							
Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type		
0.0 - 1		52	-	0.000	7470A	PPB	19 Sep 2024 12:59:04	S	Std		
0.2 - 1		828	-	0.200	7470A	PPB	19 Sep 2024 13:01:21	S	Std		
2.5 - 1		9971	-	2.500	7470A	PPB	19 Sep 2024 13:03:38	S	Std		
5.0 - 1		20318	-	5.000	7470A	PPB	19 Sep 2024 13:05:54	S	Std		
7.5 - 1		30858	-	7.500	7470A	PPB	19 Sep 2024 13:08:12	S	Std		
10.0 - 1		39135	-	10.000	7470A	PPB	19 Sep 2024 13:10:29	S	Std		
ICV34 - 1	ICV34	15917	3.9628	-	7470A	PPB	19 Sep 2024 13:18:12	U	SMPL		
ICB34 - 1	ICB34	-172	-0.0821	-	7470A	PPB	19 Sep 2024 13:20:26	U	SMPL		
CCV01 - 1	CCV01	19440	4.8486	-	7470A	PPB	19 Sep 2024 13:22:43	U	SMPL		
CCB01 - 1	CCB01	-229	-0.0964	-	7470A	PPB	19 Sep 2024 13:24:57	U	SMPL		
CRA - 1	CRA	745	0.1484	-	7470A	PPB	19 Sep 2024 13:27:14	U	SMPL		
HighStd - 1	HighStd	38630	9.6731	-	7470A	PPB	19 Sep 2024 13:29:29	U	SMPL		
ChkStd - 1	ChkStd	27372	6.8427	-	7470A	PPB	19 Sep 2024 13:31:44	U	SMPL		
PB163475BL - 1	PBW	-159	-0.0788	-	7470A	PPB	19 Sep 2024 13:34:02	U	SMPL		
PB163475BS - 1	LCSW	15962	3.9742	-	7470A	PPB	19 Sep 2024 13:39:44	U	SMPL		
P4085-05 - 1	PURGE-WATER	-56	-0.0529	-	7470A	PPB	19 Sep 2024 13:42:00	U	SMPL		
P4085-05DUP - 1	PURGE-WATERDUP	114	-0.0102	-	7470A	PPB	19 Sep 2024 13:49:41	U	SMPL		
P4085-05MS - 1	PURGE-WATERMS	15673	3.9015	-	7470A	PPB	19 Sep 2024 13:51:57	U	SMPL		
P4085-05MSD - 1	PURGE-WATERMSD	16289	4.0564	-	7470A	PPB	19 Sep 2024 13:54:14	U	SMPL		
P4097-03 - 1	FM648	43748	10.9598	-	7470A	PPB	19 Sep 2024 13:56:30	U	SMPL		
CCV02 - 1	CCV02	20081	5.0097	-	7470A	PPB	19 Sep 2024 14:01:31	U	SMPL		
CCB02 - 1	CCB02	-162	-0.0796	-	7470A	PPB	19 Sep 2024 14:03:46	U	SMPL		
PB163494BL - 1	PBW	250	0.0240	-	7470A	PPB	19 Sep 2024 14:06:03	U	SMPL		
PB163494BS - 1	LCSW	15028	3.7393	-	7470A	PPB	19 Sep 2024 14:08:18	U	SMPL		
P4030-02 - 1	WC-A-08-202409	1413	0.3164	-	7470A	PPB	19 Sep 2024 14:10:33	U	SMPL		
P4085-02 - 1	COMP-1	1084	0.2337	-	7470A	PPB	19 Sep 2024 14:12:50	U	SMPL		
P4085-04 - 1	COMP-2	1879	0.4335	-	7470A	PPB	19 Sep 2024 14:15:05	U	SMPL		
P4087-02 - 1	NB-614-COMP-01	1859	0.4285	-	7470A	PPB	19 Sep 2024 14:17:21	U	SMPL		
P4087-08 - 1	NB-614-COMP-02	1549	0.3506	-	7470A	PPB	19 Sep 2024 14:19:37	U	SMPL		
P4087-14 - 1	NB-614-COMP-03	1896	0.4378	-	7470A	PPB	19 Sep 2024 14:21:54	U	SMPL		
P4087-20 - 1	NB-614-COMP-04	2995	0.7141	-	7470A	PPB	19 Sep 2024 14:24:11	U	SMPL		
P4087-26 - 1	NB-614-COMP-05	2046	0.4755	-	7470A	PPB	19 Sep 2024 14:26:26	U	SMPL		
CCV03 - 1	CCV03	19952	4.9773	-	7470A	PPB	19 Sep 2024 14:28:41	U	SMPL		
CCB03 - 1	CCB03	-149	-0.0763	-	7470A	PPB	19 Sep 2024 14:30:57	U	SMPL		
P4087-32 - 1	NB-614-COMP-06	2337	0.5487	-	7470A	PPB	19 Sep 2024 14:33:14	U	SMPL		
P4088-02 - 1	NB-614-COMP-07	2052	0.4770	-	7470A	PPB	19 Sep 2024 14:35:29	U	SMPL		
P4088-08 - 1	NB-614-COMP-08	2012	0.4670	-	7470A	PPB	19 Sep 2024 14:37:46	U	SMPL		
P4088-14 - 1	NB-614-COMP-09	1941	0.4491	-	7470A	PPB	19 Sep 2024 14:40:01	U	SMPL		
P4089-02 - 1	365-351-294	2397	0.5638	-	7470A	PPB	19 Sep 2024 14:42:17	U	SMPL		
P4097-02 - 1	RT2299	1938	0.4484	-	7470A	PPB	19 Sep 2024 14:44:34	U	SMPL		
P4097-02DUP - 1	RT2299DUP	1884	0.4348	-	7470A	PPB	19 Sep 2024 14:46:50	U	SMPL		
P4097-02MS - 1	RT2299MS	14773	3.6752	-	7470A	PPB	19 Sep 2024 14:49:07	U	SMPL		
P4097-02MSD - 1	RT2299MSD	14753	3.6702	-	7470A	PPB	19 Sep 2024 14:51:23	U	SMPL		
CCV04 - 1	CCV04	19450	4.8511	-	7470A	PPB	19 Sep 2024 14:56:41	U	SMPL		
CCB04 - 1	CCB04	-234	-0.0977	-	7470A	PPB	19 Sep 2024 14:58:56	U	SMPL		
P4085-05LX5 - 1	P4085-05L	5	-0.0376	-	7470A	PPB	19 Sep 2024 15:01:13	U	SMPL		
P4085-05A - 1	P4085-05A	15689	3.9055	-	7470A	PPB	19 Sep 2024 15:03:28	U	SMPL		
P4097-02LX5 - 1	RT2299A	221	0.0167	-	7470A	PPB	19 Sep 2024 15:05:43	U	SMPL		
P4097-02A - 1	RT2299A	16954	4.2235	-	7470A	PPB	19 Sep 2024 15:08:00	U	SMPL		
PB163446TB - 1	LEB446	-193	-0.0874	-	7470A	PPB	19 Sep 2024 15:10:16	U	SMPL		
P4097-03DLX2 - 1		22282	5.5631	-	7470A	PPB	19 Sep 2024 15:12:33	U	SMPL		
CCV05 - 1	CCV05	19754	4.9275	-	7470A	PPB	19 Sep 2024 15:14:52	U	SMPL		
CCB05 - 1	CCB05	-195	-0.0879	-	7470A	PPB	19 Sep 2024 15:17:11	U	SMPL		

9/19/24