


LB132517, INSTRUMENT ID: CVI

245.1

Linear



A= 0.0000e+000

B= 2.6402e-004

C= -1.2483e-002

Rho= 0.9997867

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	1/D
0.0	0.000	0.004	0.004	62	0.000	62					
0.05	0.050	0.043	-0.007	210	0.0 %	210	0				-14
0.2	0.200	0.198	-0.002	797	0.0 %	797					-1
2.5	2.500	2.472	-0.028	9411	0.0 %	9411					-1
5.0	5.000	4.977	-0.023	18899	0.0 %	18899					-1
7.5	7.500	7.666	0.166	29084	0.0 %	29084					2
10.0	10.000	9.890	-0.110	37506	0.0 %	37506					-1

LB132517, INSTRUMENT ID:CV1

Method: 245.1

Operator: Admin

Date of Analysis: 19 Sep 2024 16:58:29

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50.0	62	-	0.0000	245.1	PPB	19 Sep 2024 17:20:46	S	Std
0.05 - 1	50.05	210	-	0.0500	245.1	PPB	19 Sep 2024 17:27:45	S	Std
0.2 - 1	50.20	797	-	0.2000	245.1	PPB	19 Sep 2024 17:30:03	S	Std
2.5 - 1	52.50	9411	-	2.5000	245.1	PPB	19 Sep 2024 17:32:19	S	Std
5.0 - 1	55.0	18899	-	5.0000	245.1	PPB	19 Sep 2024 17:34:35	S	Std
7.5 - 1	57.5	29084	-	7.5000	245.1	PPB	19 Sep 2024 17:36:51	S	Std
10.0 - 1	510.0	37506	-	10.0000	245.1	PPB	19 Sep 2024 17:54:29	S	Std
ICV35 - 1	ICV35	15788	4.1558	-245.1		PPB	19 Sep 2024 17:57:12	U	SMPL
ICB35 - 1	ICB35	140	0.0245	-245.1		PPB	19 Sep 2024 18:02:07	U	SMPL
CCV01 - 1	CCV01	20069	5.2861	-245.1		PPB	19 Sep 2024 18:04:21	U	SMPL
CCB01 - 1	CCB01	25	-0.0059	-245.1		PPB	19 Sep 2024 18:06:36	U	SMPL
CRA - 1	CRA	191	0.0379	-245.1		PPB	19 Sep 2024 18:08:53	U	SMPL
ChkStd - 1	ChkStd	26316	6.9354	-245.1		PPB	19 Sep 2024 18:13:23	U	SMPL
PB163478BL - 1	PBW	-67	-0.0302	-245.1		PPB	19 Sep 2024 18:15:39	U	SMPL
PB163478BS - 1	LCSW	16222	4.2704	-245.1		PPB	19 Sep 2024 18:17:57	U	SMPL
HighStd - 1	HighStd	38216	10.0772	-245.1		PPB	19 Sep 2024 18:21:22	U	SMPL
P3883-02 - 1	FALL-2024	5232	1.3689	-245.1		PPB	19 Sep 2024 18:26:42	U	SMPL
P3940-01 - 1	AST-2024-4036-PRIM	409	0.0955	-245.1		PPB	19 Sep 2024 18:28:59	U	SMPL
P3956-02 - 1	EFF-WW	760	0.1882	-245.1		PPB	19 Sep 2024 18:31:18	U	SMPL
P3956-02DUP - 1	EFF-WWDUP	567	0.1372	-245.1		PPB	19 Sep 2024 18:33:33	U	SMPL
P3956-02MS - 1	CCV02-1 EFF-WWMS CCV02	18631	4.9064	-245.1		PPB	19 Sep 2024 18:35:48	U	SMPL
CCV02 - 1	CCB02-1 CCV02-CCB02	35	-0.0032	-245.1		PPB	19 Sep 2024 18:38:03	U	SMPL
P3956-02MS - 1	EFF-WWMS	16404	4.3184	-245.1		PPB	19 Sep 2024 18:46:48	U	SMPL
P3956-02MSD - 1	EFF-WWMSD	16137	4.2479	-245.1		PPB	19 Sep 2024 18:49:03	U	SMPL
P3956-02MSD - 1	EFF-WWMSD	16066	4.2292	-245.1		PPB	19 Sep 2024 18:51:45	U	SMPL
P3956-02LX5 - 1	FALL-2024	25	-0.0059	-245.1		PPB	19 Sep 2024 18:56:16	U	SMPL
CCV03 - 1	CCV03	19646	5.1744	-245.1		PPB	19 Sep 2024 18:58:33	U	SMPL
CCB03 - 1	CCB03	-18	-0.0172	-245.1		PPB	19 Sep 2024 19:00:49	U	SMPL

SD 9/19/24