

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132525

Slope : 98.2

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/20/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	09/20/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	09/20/2024	08:19
4	ICV	1	Water	NA	NA	20.3	7.00	09/20/2024	08:22
5	CCV1	1	Water	NA	NA	20.2	2.01	09/20/2024	08:25
6	P4110-02	1	Water	NA	NA	20.9	7.25	09/20/2024	08:30
7	P4110-03	1	Water	NA	NA	20.7	7.32	09/20/2024	08:37
8	P4118-01	1	Water	NA	NA	24.2	6.17	09/20/2024	08:44
9	P4118-01DUP	1	Water	NA	NA	24.3	6.18	09/20/2024	08:45
10	CCV2	1	Water	NA	NA	20.3	12.02	09/20/2024	08:47

132525

WORKLIST(Hardcopy Internal Chain)

WorkList Name : PH P4123 WorkList ID : 183669 Department : Wet-Chemistry Date : 09-20-2024 08:00:08

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4110-02	TT-065-IDWGW-20240918	Water	pH	Cool 4 deg C	TETR06	J11	09/18/2024	9040C
P4110-03	TT-066-IDWGW-20240918	Water	pH	Cool 4 deg C	TETR06	J11	09/18/2024	9040C
P4118-01	091324-GLYCOL-DRUM	Water	pH	Cool 4 deg C	PSEG03	J12	09/19/2024	9040C

Date/Time09.20.2408:10

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time09.20.2410:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]