


LB132538, INSTRUMENT ID: CV1

7470A

Linear



A= 0.0000e+000

B= 2.4220e-004

C= -1.3515e-001

Rho= 0.9991931

Accept=Accepted

| Std ID | Conc.  | Calc. | Dev.   | Mean  | SD or %RSD | Rep 1 | Rep 2 | Rep 3 | Rep 4 | Rep 5 |
|--------|--------|-------|--------|-------|------------|-------|-------|-------|-------|-------|
| 0.0    | 0.000  | 0.004 | 0.004  | 576   | 0.000      | 576   | 0     |       |       |       |
| 0.2    | 0.200  | 0.205 | 0.005  | 1406  | 0.0 %      | 1406  |       |       |       |       |
| 2.5    | 2.500  | 2.420 | -0.080 | 10550 | 0.0 %      | 10550 |       |       |       |       |
| 5.0    | 5.000  | 4.982 | -0.018 | 21128 | 0.0 %      | 21128 |       |       |       |       |
| 7.5    | 7.500  | 7.790 | 0.290  | 32723 | 0.0 %      | 32723 |       |       |       |       |
| 10.0   | 10.000 | 9.798 | -0.202 | 41013 | 0.0 %      | 41013 |       |       |       |       |

7470A

## LB132538, INSTRUMENT ID:CV1

| Method: 7470A   |                          | Operator: Admin |         | Date of Analysis: 20 Sep 2024 17:38:34 |        |       |                      |      |      |
|-----------------|--------------------------|-----------------|---------|----------------------------------------|--------|-------|----------------------|------|------|
| Sample ID       | Extended ID              | µ Abs           | Conc.   | Std Conc                               | Method | Units | Date                 | Type | Type |
| 0.0 - 1         | 58.0                     | 576             | -       | 0.0000                                 | 7470A  | PPB   | 20 Sep 2024 17:41:19 | S    | Std  |
| 0.2 - 1         | 50.2                     | 1406            | -       | 0.2000                                 | 7470A  | PPB   | 20 Sep 2024 17:43:36 | S    | Std  |
| 2.5 - 1         | 52.5                     | 10550           | -       | 2.5000                                 | 7470A  | PPB   | 20 Sep 2024 17:45:53 | S    | Std  |
| 5.0 - 1         | 55.0                     | 21128           | -       | 5.0000                                 | 7470A  | PPB   | 20 Sep 2024 17:48:09 | S    | Std  |
| 7.5 - 1         | 57.5                     | 32723           | -       | 7.5000                                 | 7470A  | PPB   | 20 Sep 2024 17:50:26 | S    | Std  |
| 10.0 - 1        | 510.0                    | 41013           | -       | 10.0000                                | 7470A  | PPB   | 20 Sep 2024 17:52:44 | S    | Std  |
| ICV38 - 1       | ICV38                    | 16401           | 3.8371  | -                                      | 7470A  | PPB   | 20 Sep 2024 17:58:11 | U    | SMPL |
| ICB38 - 1       | ICB38                    | 21              | -0.1301 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:00:26 | U    | SMPL |
| CCV01 - 1       | CCV01                    | 21533           | 5.0801  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:02:43 | U    | SMPL |
| CCB01 - 1       | CCB01                    | -40             | -0.1448 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:04:57 | U    | SMPL |
| CRA - 1         | CRA                      | 1397            | 0.2032  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:07:14 | U    | SMPL |
| HighStd - 1     | HighStd                  | 41008           | 9.7968  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:09:28 | U    | SMPL |
| ChkStd - 1      | ChkStd                   | 30003           | 7.1315  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:11:44 | U    | SMPL |
| PB163549BL - 1  | PBW                      | 18              | -0.1308 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:14:02 | U    | SMPL |
| PB163549BS - 1  | LCSW                     | 17220           | 4.0355  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:16:19 | U    | SMPL |
| P4110-02 - 1    | TT-065-IDWGW-20240918    | 208             | -0.0848 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:18:35 | U    | SMPL |
| P4110-02DUP - 1 | TT-065-IDWGW-20240918DUP | 340             | -0.0528 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:20:53 | U    | SMPL |
| P4110-02MS - 1  | TT-065-IDWGW-20240918MS  | 15228           | 3.5530  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:23:10 | U    | SMPL |
| P4110-02MSD - 1 | TT-065-IDWGW-20240918MSD | 16382           | 3.8325  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:25:27 | U    | SMPL |
| P4110-03 - 1    | TT-066-IDWGW-20240918    | 222             | -0.0814 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:27:45 | U    | SMPL |
| CCV02 - 1       | CCV02                    | 21136           | 4.9839  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:30:01 | U    | SMPL |
| CCB02 - 1       | CCB02                    | 288             | -0.0654 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:32:17 | U    | SMPL |
| P4110-02LX5 - 1 |                          | 127             | -0.1044 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:34:35 | U    | SMPL |
| P4110-02A - 1   |                          | 16324           | 3.8185  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:36:50 | U    | SMPL |
| CCV03 - 1       | CCV03                    | 21296           | 5.0227  | -                                      | 7470A  | PPB   | 20 Sep 2024 18:39:05 | U    | SMPL |
| CCB03 - 1       | CCB03                    | 300             | -0.0625 | -                                      | 7470A  | PPB   | 20 Sep 2024 18:41:22 | U    | SMPL |