


SFAM01.1



Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5
0.00	0.000	-0.022	-0.022	343	0.000	343				
0.05	0.050					442	0			
0.20	0.200	0.180	-0.020	1237	0.0 %	1237				
2.50	2.500	2.360	-0.140	10921	0.0 %	10921				
5.00	5.000	5.192	0.192	23497	0.0 %	23497				
7.50	7.500	7.801	0.301	35085	0.0 %	35085				
10.0	10.000	9.690	-0.310	43476	0.0 %	43476				

ANALYSIS DATE:092124, INSTRUMENT ID:CV1, LB132540

Sample ID	Extended II μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
0 S0	343	-		0 SFAM01.1	PPB	9/21/2024 15:27	Std
0.2 S01	1237	-		0.2 SFAM01.1	PPB	9/21/2024 15:29	Std
2.5 S02	10921	-		2.5 SFAM01.1	PPB	9/21/2024 15:36	Std
5 S03	23497	-		5 SFAM01.1	PPB	9/21/2024 15:38	Std
7.5 S04	35085	-		7.5 SFAM01.1	PPB	9/21/2024 15:40	Std
10 S05	43476	-		10 SFAM01.1	PPB	9/21/2024 15:43	Std
ICV001	ICV001	19837	4.3675	SFAM01.1	PPB	9/21/2024 15:46	SMPL
ICB001	ICB001	372	-0.0151	SFAM01.1	PPB	9/21/2024 15:50	SMPL
CCV001	CCV001	22392	4.9427	SFAM01.1	PPB	9/21/2024 15:53	SMPL
CCB001	CCB001	205	-0.0527	SFAM01.1	PPB	9/21/2024 15:55	SMPL
PB163583E	PB163583E	333	-0.0239	SFAM01.1	PPB	9/21/2024 15:57	SMPL
P3931-01	YEOP7	7621	1.617	SFAM01.1	PPB	9/21/2024 15:59	SMPL
P3931-02	YE0Q0	4792	0.9801	SFAM01.1	PPB	9/21/2024 16:02	SMPL
P3931-03	YEOR3	8487	1.812	SFAM01.1	PPB	9/21/2024 16:04	SMPL
P3931-04	YEOR4	7331	1.5517	SFAM01.1	PPB	9/21/2024 16:06	SMPL
P3931-05	YEOR5	2683	0.5052	SFAM01.1	PPB	9/21/2024 16:08	SMPL
P3931-06	YEOR6	2704	0.51	SFAM01.1	PPB	9/21/2024 16:11	SMPL
P3931-07	YEOR7	2767	0.5241	SFAM01.1	PPB	9/21/2024 16:13	SMPL
P3931-08	YEOR8	2229	0.403	SFAM01.1	PPB	9/21/2024 16:15	SMPL
P3931-09	YEOR8D	2065	0.3661	SFAM01.1	PPB	9/21/2024 16:17	SMPL
P3931-10	YEOR8S	7941	1.6891	SFAM01.1	PPB	9/21/2024 16:20	SMPL
P3931-11	YEOR9	2920	0.5586	SFAM01.1	PPB	9/21/2024 16:22	SMPL
P3931-12	YE0S0	4668	0.9522	SFAM01.1	PPB	9/21/2024 16:24	SMPL
P3931-13	YE0S2	3278	0.6392	SFAM01.1	PPB	9/21/2024 16:26	SMPL
P3931-14	YE0Q9	6116	1.2782	SFAM01.1	PPB	9/21/2024 16:29	SMPL
P3931-15	YEOR1	2893	0.5525	SFAM01.1	PPB	9/21/2024 16:31	SMPL
P3931-16	YEOR2	2301	0.4192	SFAM01.1	PPB	9/21/2024 16:33	SMPL
P3921-01	YEOP2	3894	0.7779	SFAM01.1	PPB	9/21/2024 16:38	SMPL
P3921-02	YEOP3	32786	7.283	SFAM01.1	PPB	9/21/2024 16:40	SMPL
CCV002	CCV002	22637	4.9979	SFAM01.1	PPB	9/21/2024 16:42	SMPL
CCB002	CCB002	171	-0.0604	SFAM01.1	PPB	9/21/2024 16:45	SMPL
P3921-03	YEOP4	6536	1.3727	SFAM01.1	PPB	9/21/2024 16:47	SMPL
P3921-04	YEOP8	13475	2.9351	SFAM01.1	PPB	9/21/2024 16:49	SMPL
P3921-05	YEOP9	13173	2.8671	SFAM01.1	PPB	9/21/2024 16:51	SMPL
P3921-06	YE0Q1	8472	1.8086	SFAM01.1	PPB	9/21/2024 16:54	SMPL
P3921-07	YE0Q1D	7854	1.6695	SFAM01.1	PPB	9/21/2024 16:56	SMPL
P3921-08	YE0Q1S	16050	3.5148	SFAM01.1	PPB	9/21/2024 16:58	SMPL
P3921-09	YE0Q7	15329	3.3525	SFAM01.1	PPB	9/21/2024 17:00	SMPL
P3921-10	YE0Q8	14863	3.2476	SFAM01.1	PPB	9/21/2024 17:03	SMPL
P3921-11	YEOR0	14054	3.0654	SFAM01.1	PPB	9/21/2024 17:05	SMPL
P3921-12	YE0S1	3118	0.6032	SFAM01.1	PPB	9/21/2024 17:07	SMPL
P3921-13	YE0S3	13001	2.8283	SFAM01.1	PPB	9/21/2024 17:10	SMPL
P3921-15	YEOP1	4379	0.8871	SFAM01.1	PPB	9/21/2024 17:12	SMPL
P3921-16	YEOP5	9045	1.9376	SFAM01.1	PPB	9/21/2024 17:14	SMPL
P3921-17	YEOP6	6777	1.427	SFAM01.1	PPB	9/21/2024 17:16	SMPL
P3921-18	YE0Q2	3599	0.7115	SFAM01.1	PPB	9/21/2024 17:19	SMPL

P3921-19	YE0Q3	2666	0.5014 -	SFAM01.1 PPB	9/21/2024 17:21 SMPL
P3921-20	YE0Q4	4108	0.8261 -	SFAM01.1 PPB	9/21/2024 17:23 SMPL
P3921-21	YE0Q5	3030	0.5834 -	SFAM01.1 PPB	9/21/2024 17:25 SMPL
P3921-22	YE0Q6	8022	1.7073 -	SFAM01.1 PPB	9/21/2024 17:28 SMPL
CCV003	CCV003	23105	5.1033 -	SFAM01.1 PPB	9/21/2024 17:30 SMPL
CCB003	CCB103	167	-0.0613 -	SFAM01.1 PPB	9/21/2024 17:32 SMPL
PB163582E	PBW582	267	-0.0387 -	SFAM01.1 PPB	9/21/2024 17:35 SMPL
CCV004	CCV004	22863	5.0488 -	SFAM01.1 PPB	9/21/2024 17:37 SMPL
CCB004	CCB004	162	-0.0624 -	SFAM01.1 PPB	9/21/2024 17:39 SMPL