

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132544

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/23/2024	13:25
2	P4119-01		1	11	100.00	>212.0	>212.0	09/23/2024	13:50
3	P4119-01DUP		1	11	100.00	>212.0	>212.0	09/23/2024	14:50
4	P4121-03		1	4	7.00	44.6	43.9	09/23/2024	15:20

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132544

WorkList Name : fp s9-20 WorkList ID : 183688 Department : Wet-Chemistry Date : 09-20-2024 08:55:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4119-01	TRANSFOMER-OIL	Solid	Flash Point	Cool 4 deg C	PSEG03	J12	09/19/2024	1010B
P4121-03	ELZ-24-00090	Solid	Flash Point	Cool 4 deg C	PSEG03	J13	09/19/2024	1010B

Date/Time 09/23/2024 13:15
Raw Sample Received by: RM wcy
Raw Sample Relinquished by: D Wacey

Date/Time 09/23/2024 16:10
Raw Sample Received by: RM wcy
Raw Sample Relinquished by: RM wcy