

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB132561

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.5

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/24/2024	08:11
2	CAL2	1	Water	NA	NA	20.3	7.00	09/24/2024	08:12
3	CAL3	1	Water	NA	NA	20.3	10.02	09/24/2024	08:15
4	ICV	1	Water	NA	NA	20.3	7.00	09/24/2024	08:20
5	CCV1	1	Water	NA	NA	20.4	2.02	09/24/2024	08:37
6	P4156-01	1	Water	NA	NA	23.9	8.11	09/24/2024	08:42
7	P4156-01DUP	1	Water	NA	NA	23.8	8.12	09/24/2024	08:43
8	P4157-01	1	Water	NA	NA	23.7	8.05	09/24/2024	08:47
9	CCV2	1	Water	NA	NA	20.3	12.02	09/24/2024	08:52

WORKLIST(Hardcopy Internal Chain)

132561

WorkList Name : ph p3453 water WorkList ID : 183750 Department : Wet-Chemistry Date : 09-24-2024 07:57:50

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4156-01	B GAS-CLI-071024-C	Water	pH	Cool 4 deg C	PSEG03	J31	09/20/2024	9040C
P4157-01	A- 34688	Water	pH	Cool 4 deg C	PSEG03	J31	09/20/2024	9040C

Date/Time 09/24/24 08:00
Raw Sample Received by: *JP WC*
Raw Sample Relinquished by: *JP WC*

Date/Time 09/24/24 10:00
Raw Sample Received by: *JP WC*
Raw Sample Relinquished by: *JP WC*