

LB132563

7470A

INSTRUMENT ID: CY1

Linear

 μ Abs.:

43622

A= 0.0000e+000

B= 2.2398e-004

C= -1.6742e-002

Rho= 0.9989968

Accept=Accepted

Max. Conc: — 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	-0.007	-0.007	43	0.000	43	0				-
0.2	0.200	0.193	-0.007	938	0.0 %	938					-4
2.5	2.500	2.397	-0.103	10775	0.0 %	10775					-4
5.0	5.000	5.067	0.067	22696	0.0 %	22696					1
7.5	7.500	7.797	0.297	34884	0.0 %	34884					4
10.0	10.000	9.754	-0.246	43622	0.0 %	43622					-2

LB132563 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 24 Sep 2024 10:18:41

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SO	43	-	0.0000	7470A	PPB	24 Sep 2024 10:19:43	S	Std
0.2 - 1	SO-2	938	-	0.2000	7470A	PPB	24 Sep 2024 10:22:00	S	Std
2.5 - 1	52.5	10775	-	2.5000	7470A	PPB	24 Sep 2024 10:24:16	S	Std
5.0 - 1	85	22696	-	5.0000	7470A	PPB	24 Sep 2024 10:26:33	S	Std
7.5 - 1	57.5	34884	-	7.5000	7470A	PPB	24 Sep 2024 10:31:26	S	Std
10.0 - 1	51V	43622	-	10.0000	7470A	PPB	24 Sep 2024 10:33:44	S	Std
ICV01 - 1	ICV01	17411	3.8830	-7470A	PPB	PPB	24 Sep 2024 10:36:54	U	SMPL
ICB01 - 1	ICB01	-218	-0.0656	-7470A	PPB	PPB	24 Sep 2024 10:39:08	U	SMPL
CCV01 - 1	CCV01	23302	5.2025	-7470A	PPB	PPB	24 Sep 2024 10:41:25	U	SMPL
CCB01 - 1	CCB01	-206	-0.0629	-7470A	PPB	PPB	24 Sep 2024 10:43:40	U	SMPL
CRA - 1	CRA	925	0.1904	-7470A	PPB	PPB	24 Sep 2024 10:45:57	U	SMPL
HighStd - 1	HighStd	43758	9.7842	-7470A	PPB	PPB	24 Sep 2024 10:48:12	U	SMPL
ChkStd - 1	ChkStd	32438	7.2488	-7470A	PPB	PPB	24 Sep 2024 10:50:27	U	SMPL
PB163624BL - 1	PBW	-83	-0.0353	-7470A	PPB	PPB	24 Sep 2024 10:55:31	U	SMPL
PB163624BS - 1	LCSW	20811	4.6445	-7470A	PPB	PPB	24 Sep 2024 10:57:47	U	SMPL
P4138-02 - 1	LIBERTY-AVE-TP-COMP	-208	-0.0633	-7470A	PPB	PPB	24 Sep 2024 11:00:03	U	SMPL
P4140-01 - 1	MOO-23-00305	-20	-0.0212	-7470A	PPB	PPB	24 Sep 2024 11:02:21	U	SMPL
P4144-02 - 1	1-3-SBG	-1	-0.0170	-7470A	PPB	PPB	24 Sep 2024 11:04:37	U	SMPL
P4144-04 - 1	4-6-SBG	20	-0.0123	-7470A	PPB	PPB	24 Sep 2024 11:06:54	U	SMPL
P4144-04DUP - 1	4-6-SBGDUP	-81	-0.0349	-7470A	PPB	PPB	24 Sep 2024 11:09:09	U	SMPL
CCV02 - 1	CCV02	22692	5.0658	-7470A	PPB	PPB	24 Sep 2024 11:11:25	U	SMPL
CCB02 - 1	CCB02	-263	-0.0756	-7470A	PPB	PPB	24 Sep 2024 11:13:40	U	SMPL
P4144-04MS - 1	4-6-SBGMS	20884	4.6609	-7470A	PPB	PPB	24 Sep 2024 11:15:56	U	SMPL
P4144-04MSD - 1	4-6-SBGMSD	19280	4.3016	-7470A	PPB	PPB	24 Sep 2024 11:18:12	U	SMPL
PB163625BL - 1	PBW	109	0.0077	-7470A	PPB	PPB	24 Sep 2024 11:23:18	U	SMPL
PB163625BS - 1	LCSW	18034	4.0225	-7470A	PPB	PPB	24 Sep 2024 11:25:32	U	SMPL
P4123-04 - 1	910-A	-129	-0.0456	-7470A	PPB	PPB	24 Sep 2024 11:27:48	U	SMPL
P4140-02 - 1	306-307	-144	-0.0490	-7470A	PPB	PPB	24 Sep 2024 11:30:06	U	SMPL
P4140-02DUP - 1	306-307DUP	-114	-0.0423	-7470A	PPB	PPB	24 Sep 2024 11:32:23	U	SMPL
P4140-02MS - 1	306-307MS	19693	4.3941	-7470A	PPB	PPB	24 Sep 2024 11:34:39	U	SMPL
P4140-02MSD - 1	306-307MSD	18876	4.2111	-7470A	PPB	PPB	24 Sep 2024 11:36:56	U	SMPL
CCV03 - 1	CCV03	22751	5.0791	-7470A	PPB	PPB	24 Sep 2024 11:39:13	U	SMPL
CCB03 - 1	CCB03	-271	-0.0774	-7470A	PPB	PPB	24 Sep 2024 11:41:30	U	SMPL
PB163591TB - 1		13	-0.0138	-7470A	PPB	PPB	24 Sep 2024 11:43:48	U	SMPL
PB163590TB - 1		12	-0.0141	-7470A	PPB	PPB	24 Sep 2024 11:46:03	U	SMPL
P4144-04LX5 - 1		138	0.0142	-7470A	PPB	PPB	24 Sep 2024 11:48:19	U	SMPL
P4144-04A - 1		16976	3.7856	-7470A	PPB	PPB	24 Sep 2024 11:50:34	U	SMPL
P4140-02LX5 - 1		-183	-0.0577	-7470A	PPB	PPB	24 Sep 2024 11:52:49	U	SMPL
P4140-02A - 1		18625	4.1549	-7470A	PPB	PPB	24 Sep 2024 11:55:08	U	SMPL
CCV04 - 1	CCV04	22859	5.1032	-7470A	PPB	PPB	24 Sep 2024 11:59:59	U	SMPL
CCB04 - 1	CCB04	-251	-0.0730	-7470A	PPB	PPB	24 Sep 2024 12:02:15	U	SMPL