

LB 132567

7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51096

A= 0.0000e+000

B= 1.9675e-004

C= 1.3113e-002

Rho= 0.9999221

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	0/0))
0.0	0.000	0.012	0.012	-7	0.000	-7					
0.2	0.200	0.219	0.019	1047	0.0 %	1047					
2.5	2.500	2.490	-0.010	12587	0.0 %	12587					
5.0	5.000	5.001	0.001	25352	0.0 %	25352					
7.5	7.500	7.412	-0.088	37606	0.0 %	37606					
10.0	10.000	10.066	0.066	51096	0.0 %	51096					

10001

LB132567 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 24 Sep 2024 12:30:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	-7	-	0.000	7471B	PPB	24 Sep 2024 12:59:56	S	Std
0.2 - 1	50.2	1047	-	0.200	7471B	PPB	24 Sep 2024 13:02:12	S	Std
2.5 - 1	52.5	12587	-	2.500	7471B	PPB	24 Sep 2024 13:04:29	S	Std
5.0 - 1	55	25352	-	5.000	7471B	PPB	24 Sep 2024 13:06:46	S	Std
7.5 - 1	57.5	37606	-	7.500	7471B	PPB	24 Sep 2024 13:09:04	S	Std
10.0 - 1	510	51096	-	10.000	7471B	PPB	24 Sep 2024 13:11:25	S	Std
ICV02 - 1	ICV02	20002	3.9485	-	7471B	PPB	24 Sep 2024 13:14:20	U	SMPL
ICB02 - 1	ICB02	-213	-0.0288	-	7471B	PPB	24 Sep 2024 13:16:36	U	SMPL
CCV05 - 1	CCV05	25768	5.0830	-	7471B	PPB	24 Sep 2024 13:22:11	U	SMPL
CCB05 - 1	CCB05	-345	-0.0548	-	7471B	PPB	24 Sep 2024 13:24:25	U	SMPL
CRA - 1	CRA	1014	0.2126	-	7471B	PPB	24 Sep 2024 13:26:45	U	SMPL
HighStd - 1	HighStd	51471	10.1401	-	7471B	PPB	24 Sep 2024 13:29:00	U	SMPL
ChkStd - 1	ChkStd	36179	7.1314	-	7471B	PPB	24 Sep 2024 13:31:15	U	SMPL
PB163626BL - 1	PBS	-405	-0.0666	-	7471B	PPB	24 Sep 2024 13:33:40	U	SMPL
PB163626BS - 1	LCSS	21991	4.3399	-	7471B	PPB	24 Sep 2024 13:36:02	U	SMPL
P4138-01 - 1	LIBERTY-AVE-TP-COMP	12523	2.4770	-	7471B	PPB	24 Sep 2024 13:38:18	U	SMPL
P4142-01 - 1	OR-02-092024	3649	0.7311	-	7471B	PPB	24 Sep 2024 13:40:37	U	SMPL
P4144-01 - 1	1-3-SBG	8719	1.7286	-	7471B	PPB	24 Sep 2024 13:42:57	U	SMPL
P4144-03 - 1	4-6-SBG	4838	0.9650	-	7471B	PPB	24 Sep 2024 13:45:16	U	SMPL
P4146-01 - 1	CL-02-092024	1011	0.2120	-	7471B	PPB	24 Sep 2024 13:47:32	U	SMPL
CCV06 - 1	CCV06	24087	4.7522	-	7471B	PPB	24 Sep 2024 13:49:52	U	SMPL
CCB06 - 1	CCB06	-130	-0.0125	-	7471B	PPB	24 Sep 2024 13:52:07	U	SMPL
P4148-01 - 1	STOCK-PILE	1718	0.3511	-	7471B	PPB	24 Sep 2024 13:54:26	U	SMPL
P4148-01DUP - 1	STOCK-PILEDUP	1507	0.3096	-	7471B	PPB	24 Sep 2024 13:56:40	U	SMPL
P4148-01MS - 1	STOCK-PILEMS	24086	4.7520	-	7471B	PPB	24 Sep 2024 13:58:56	U	SMPL
P4148-01MSD - 1	STOCK-PILEMSD	23490	4.6348	-	7471B	PPB	24 Sep 2024 14:01:12	U	SMPL
P4148-01LX5 - 1		248	0.0619	-	7471B	PPB	24 Sep 2024 14:05:32	U	SMPL
P4148-01A - 1		24157	4.7660	-	7471B	PPB	24 Sep 2024 14:07:48	U	SMPL
CCV07 - 1	CCV07	25210	4.9732	-	7471B	PPB	24 Sep 2024 14:10:05	U	SMPL
CCB07 - 1	CCB07	-137	-0.0138	-	7471B	PPB	24 Sep 2024 14:12:25	U	SMPL