

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132568

Slope : 98.5

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/24/2024	08:15
2	CAL2	1	Water	NA	NA	20.3	7.00	09/24/2024	08:16
3	CAL3	1	Water	NA	NA	20.3	10.02	09/24/2024	08:19
4	ICV	1	Water	NA	NA	20.3	7.02	09/24/2024	08:20
5	CCV1	1	Water	NA	NA	20.3	2.01	09/24/2024	08:25
6	P4153-01	1	Solid	20.02	20	20.8	5.62	09/24/2024	08:37
7	P4153-01DUP	1	Solid	20.03	20	20.9	5.63	09/24/2024	08:38
8	CCV2	1	Water	NA	NA	20.3	12.02	09/24/2024	08:40

WORKLIST(Hardcopy Internal Chain)

WorkList Name : ph s p4153 WorkList ID : 183752 Department : Wet-Chemistry Date : 09-24-2024 08:06:10

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4153-01	1577	Solid	pH	Cool 4 deg C	PSEG03	J33	09/23/2024	9045D

Date/Time 09/24/24 08:10
Raw Sample Received by: jd wcc
Raw Sample Relinquished by: RM SM

Date/Time 09/24/24 12:00
Raw Sample Received by: RM SM
Raw Sample Relinquished by: jd wcc