

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132573

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	09/24/2024	13:15
2	P4118-01		1	11	100.00	>212.0	>212.0	09/24/2024	13:45
3	P4118-01DUP		1	11	100.00	>212.0	>212.0	09/24/2024	14:15
4	P4123-03		1	12	100.00	>212.0	>212.0	09/24/2024	14:45
5	P4140-02		1	14	100.00	>212.0	>212.0	09/24/2024	15:15
6	P4156-01		1	11	100.00	>212.0	>212.0	09/24/2024	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66132573

WorkList Name : fp-w9-20 WorkList ID : 183689 Department : Wet-Chemistry Date : 09-20-2024 08:56:33

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4118-01	091324-GLYCOL-DRUM	Water	Flash Point	Cool 4 deg C	PSEG03	J12	09/19/2024	1010B
P4123-03	ORA-1895	Water	Flash Point	Cool 4 deg C	PSEG03	J21	09/19/2024	1010B
P4140-02	306-307	Water	Flash Point	Cool 4 deg C	PSEG03	J31	09/20/2024	1010B
P4156-01	GAS-CLI-071024-C	Water	Flash Point	Cool 4 deg C	PSEG03	J31	09/20/2024	1010B

Date/Time 09/20/2024 13:05
 Raw Sample Received by: RM CWCJ
 Raw Sample Relinquished by: JAWOC

Date/Time 09/24/2024 16:20
 Raw Sample Received by: JAWOC
 Raw Sample Relinquished by: RM CWCJ