

LB132575


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

45299



A= 0.0000e+000

B= 2.2244e-004 slope

C= 5.1641e-002 y-intercept

Rho= 0.9997388

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	6/b
0.00	0.000	0.058	0.058	29	0.000	29					=
0.05	0.050					442	0				23
0.20	0.200	0.245	0.045	869	0.0 %	869					- 2
2.50	2.500	2.448	-0.052	10771	0.0 %	10771					- 1
5.00	5.000	4.944	-0.056	21996	0.0 %	21996					- 2
7.50	7.500	7.377	-0.123	32933	0.0 %	32933					1
10.0	10.000	10.128	0.128	45299	0.0 %	45299					

LB132575

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	29	-		0 SFAM01.1	PPB	9/24/2024 15:27	Std
	0.2 S01	869	-		0.2 SFAM01.1	PPB	9/24/2024 15:30	Std
	2.5 S02	10771	-		2.5 SFAM01.1	PPB	9/24/2024 15:32	Std
	5 S03	21996	-		5 SFAM01.1	PPB	9/24/2024 15:34	Std
	7.5 S04	32933	-		7.5 SFAM01.1	PPB	9/24/2024 15:36	Std
	10 S05	45299	-		10 SFAM01.1	PPB	9/24/2024 15:41	Std
ICV003	ICV003	16652	3.7557 -		SFAM01.1	PPB	9/24/2024 15:44	SMPL
ICB003	ICB003	-239	-0.0015 -		SFAM01.1	PPB	9/24/2024 15:46	SMPL
CCV008	CCV008	22640	5.0876 -		SFAM01.1	PPB	9/24/2024 15:49	SMPL
CCB008	CCB008	-254	-0.0049 -		SFAM01.1	PPB	9/24/2024 15:51	SMPL
PB163629BL	PBW629	47	0.0621 -		SFAM01.1	PPB	9/24/2024 15:53	SMPL
P4004-01	MH2A01	3710	0.8769 -		SFAM01.1	PPB	9/24/2024 15:55	SMPL
P4004-02	MH2A02	1132	0.3034 -		SFAM01.1	PPB	9/24/2024 15:58	SMPL
P4004-03	MH2A03	2259	0.5541 -		SFAM01.1	PPB	9/24/2024 16:00	SMPL
P4004-04	MH2A04	1108	0.2981 -		SFAM01.1	PPB	9/24/2024 16:02	SMPL
P4004-05	MH2A05	474	0.1571 -		SFAM01.1	PPB	9/24/2024 16:05	SMPL
P4004-06	MH2A06	733	0.2147 -		SFAM01.1	PPB	9/24/2024 16:07	SMPL
P4004-07	MH2A07	913	0.2547 -		SFAM01.1	PPB	9/24/2024 16:09	SMPL
P4004-08	MH2A08	171	0.0897 -		SFAM01.1	PPB	9/24/2024 16:11	SMPL
P4004-09	MH2A09	289	0.1159 -		SFAM01.1	PPB	9/24/2024 16:14	SMPL
P4004-10	MH2A10	1644	0.4173 -		SFAM01.1	PPB	9/24/2024 16:16	SMPL
P4004-11	MH2A11	1699	0.4296 -		SFAM01.1	PPB	9/24/2024 16:18	SMPL
P4004-12	MH2A12	724	0.2127 -		SFAM01.1	PPB	9/24/2024 16:20	SMPL
P4004-13	MH2A13	2441	0.5946 -		SFAM01.1	PPB	9/24/2024 16:23	SMPL
P4004-14	MH2A14	206	0.0975 -		SFAM01.1	PPB	9/24/2024 16:25	SMPL
P4004-15	MH2A15	1052	0.2856 -		SFAM01.1	PPB	9/24/2024 16:27	SMPL
P4004-16	MH2A16	2313	0.5661 -		SFAM01.1	PPB	9/24/2024 16:29	SMPL
P4004-17	MH2A17	59	0.0648 -		SFAM01.1	PPB	9/24/2024 16:32	SMPL
P4004-18	MH2A18	483	0.1591 -		SFAM01.1	PPB	9/24/2024 16:34	SMPL
P4004-19	MH2A19	246	0.1064 -		SFAM01.1	PPB	9/24/2024 16:36	SMPL
CCV009	CCV009	21899	4.9228 -		SFAM01.1	PPB	9/24/2024 16:39	SMPL
CCB009	CCB009	-51	0.0403 -		SFAM01.1	PPB	9/24/2024 16:41	SMPL
P4004-20	MH2A19D	248	0.1068 -		SFAM01.1	PPB	9/24/2024 16:48	SMPL
P4004-21	MH2A19S	5340	1.2395 -		SFAM01.1	PPB	9/24/2024 16:50	SMPL
PB163630BL	PBW630	0	0.0516 -		SFAM01.1	PPB	9/24/2024 16:52	SMPL
P4005-01	MH2A20	1124	0.3017 -		SFAM01.1	PPB	9/24/2024 16:55	SMPL
P4005-02	MH2A21	42	0.061 -		SFAM01.1	PPB	9/24/2024 16:57	SMPL
P4005-03	MH2A21D	43	0.0612 -		SFAM01.1	PPB	9/24/2024 16:59	SMPL
P4005-04	MH2A21S	5640	1.3062 -		SFAM01.1	PPB	9/24/2024 17:01	SMPL
PB163578TB	LEB578	110	0.0761 -		SFAM01.1	PPB	9/24/2024 17:04	SMPL
PB163579TB	LEB579	36	0.0596 -		SFAM01.1	PPB	9/24/2024 17:06	SMPL
CCV010	CCV010	22019	4.9495 -		SFAM01.1	PPB	9/24/2024 17:08	SMPL
CCB010	CCB010	-2	0.0512 -		SFAM01.1	PPB	9/24/2024 17:11	SMPL