

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: Iwona

Parameter: Flash Point

Supervisor Review By: jignesh

Run Number: LB132577

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flashpoint

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	09/26/2024	11:00
2	P4157-01		1	12	100.00	>212.0	>212.0	09/26/2024	11:30
3	P4157-01DUP		1	11	100.00	>212.0	>212.0	09/26/2024	12:00
4	P4166-01		1	12	100.00	>212.0	>212.0	09/26/2024	12:30
5	P4172-01		1	12	100.00	>212.0	>212.0	09/26/2024	13:00
6	P4178-06		1	12	100.00	>212.0	>212.0	09/26/2024	13:30

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132577

WorkList Name : FP-9-25

WorkList ID : 183796

Department : Wet-Chemistry

Date : 09-25-2024 09:14:19

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4157-01	34688	Water	Flash Point	Cool 4 deg C	PSEG03	J31	09/20/2024	1010B
P4166-01	50423	Water	Flash Point	Cool 4 deg C	PSEG03	J31	09/24/2024	1010B
P4172-01	FMI120	Water	Flash Point	Cool 4 deg C	PSEG03	J41	09/24/2024	1010B
P4178-06	3116	Water	Flash Point	Cool 4 deg C	PSEG03	J41	09/24/2024	1010B

Date/Time 09/26/24 09:30
Raw Sample Received by: 121001
Raw Sample Relinquished by: 181001

Date/Time 09/26/24 14:10
Raw Sample Received by: 121001
Raw Sample Relinquished by: 121001