

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB132580

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/25/2024	08:30
2	CAL2	1	Water	NA	NA	20.3	7.00	09/25/2024	08:31
3	CAL3	1	Water	NA	NA	20.3	10.02	09/25/2024	08:33
4	ICV	1	Water	NA	NA	20.3	7.00	09/25/2024	08:35
5	CCV1	1	Water	NA	NA	20.3	2.01	09/25/2024	08:37
6	P4166-01	1	Water	NA	NA	20.1	7.60	09/25/2024	08:45
7	P4166-01DUP	1	Water	NA	NA	20.2	7.61	09/25/2024	08:46
8	P4172-01	1	Water	NA	NA	22.6	7.01	09/25/2024	08:50
9	P4178-06	1	Water	NA	NA	20.9	7.70	09/25/2024	08:55
10	CCV2	1	Water	NA	NA	20.3	12.02	09/25/2024	09:00

WORKLIST(Hardcopy Internal Chain)

132580

WorkList Name : ph 04172

WorkList ID : 183777

Department : Wet-Chemistry

Date : 09-25-2024 08:22:22

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4166-01	50423	Water	pH	Cool 4 deg C	PSEG03	J31	09/24/2024	9040C
P4172-01	FMI120	Water	pH	Cool 4 deg C	PSEG03	J41	09/24/2024	9040C
P4178-06	3116	Water	pH	Cool 4 deg C	PSEG03	J41	09/24/2024	9040C

Date/Time 09/25/24 08:25

Raw Sample Received by: 70 60C1

Raw Sample Relinquished by: 70 60C1

Date/Time 09/25/24

Raw Sample Received by: 70 60C1

Raw Sample Relinquished by: 70 60C1