

LB132607


7470A

INSTRUMENT ID: CV1

Linear

 μ Abs.:

44654



A= 0.0000e+000

B= 2.2115e-004

C= 6.8937e-002

Rho= 0.9985307

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.0	0.000	0.093	0.093	108	0.000	108					1
0.2	0.200	0.235	0.035	752	0.0 %	752					18
2.5	2.500	2.519	0.019	11079	0.0 %	11079					1
5.0	5.000	4.621	-0.379	20582	0.0 %	20582					8
7.5	7.500	7.788	0.288	34902	0.0 %	34902					4
10.0	10.000	9.944	-0.056	44654	0.0 %	44654					1

LB132607 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 26 Sep 2024 10:09:07

Sample ID	Extended ID	μ Abs.	Conc.	Std Cond	Method	Units	Date	Type	Type
0.0 - 1	50	108	-	0.0000	7470A	PPB	26 Sep 2024 10:16:10	S	Std
0.2 - 1	50.2	752	-	0.2000	7470A	PPB	26 Sep 2024 10:18:26	S	Std
2.5 - 1	52.5	11079	-	2.5000	7470A	PPB	26 Sep 2024 10:20:43	S	Std
5.0 - 1	55	20582	-	5.0000	7470A	PPB	26 Sep 2024 10:23:00	S	Std
7.5 - 1	57.5	34902	-	7.5000	7470A	PPB	26 Sep 2024 10:25:17	S	Std
10.0 - 1	510	44654	-	10.0000	7470A	PPB	26 Sep 2024 10:27:34	S	Std
ICV07 - 1	ICV07	16795	3.7832	-	7470A	PPB	26 Sep 2024 10:30:47	U	SMPL
ICB07 - 1	ICB07	-40	0.0601	-	7470A	PPB	26 Sep 2024 10:33:01	U	SMPL
CCV21 - 1	CCV21	23627	5.2942	-	7470A	PPB	26 Sep 2024 10:42:23	U	SMPL
CCB21 - 1	CCB21	-241	0.0156	-	7470A	PPB	26 Sep 2024 10:44:40	U	SMPL
CRA - 1	CRA	704	0.2246	-	7470A	PPB	26 Sep 2024 10:46:57	U	SMPL
HighStd - 1	HighStd	44376	9.8829	-	7470A	PPB	26 Sep 2024 10:49:11	U	SMPL
ChkStd - 1	ChkStd	30821	6.8852	-	7470A	PPB	26 Sep 2024 10:51:27	U	SMPL
P4148-02 - 1	STOCK-PILE	-335	-0.0052	-	7470A	PPB	26 Sep 2024 10:53:45	U	SMPL
P4148-02DUP - 1	STOCK-PILEDUP	11	0.0714	-	7470A	PPB	26 Sep 2024 10:56:02	U	SMPL
P4148-02MS - 1	STOCK-PILEMS	18145	4.0818	-	7470A	PPB	26 Sep 2024 11:04:03	U	SMPL
P4148-02MSD - 1	STOCK-PILEMSD	18817	4.2304	-	7470A	PPB	26 Sep 2024 11:06:19	U	SMPL
P4169-01 - 1	D3652-55	-162	0.0331	-	7470A	PPB	26 Sep 2024 11:08:37	U	SMPL
P4172-03 - 1	OIL-TOTE-0924	-4	0.0681	-	7470A	PPB	26 Sep 2024 11:10:55	U	SMPL
P4178-01 - 1	3114	197	0.1125	-	7470A	PPB	26 Sep 2024 11:13:10	U	SMPL
CCV22 - 1	CCV22	21661	4.8594	-	7470A	PPB	26 Sep 2024 11:18:08	U	SMPL
CCB22 - 1	CCB22	-197	0.0254	-	7470A	PPB	26 Sep 2024 11:20:23	U	SMPL
P4178-02 - 1	3129	58	0.0818	-	7470A	PPB	26 Sep 2024 11:22:40	U	SMPL
P4178-03 - 1	3022-3023	101	0.0913	-	7470A	PPB	26 Sep 2024 11:24:55	U	SMPL
PB163673BL - 1	PBW	32	0.0760	-	7470A	PPB	26 Sep 2024 11:27:10	U	SMPL
PB163673BS - 1	LCSW	17932	4.0347	-	7470A	PPB	26 Sep 2024 11:29:26	U	SMPL
P4178-06 - 1	3116	67	0.0838	-	7470A	PPB	26 Sep 2024 11:31:41	U	SMPL
P4178-06DUP - 1	3116DUP	211	0.1156	-	7470A	PPB	26 Sep 2024 11:33:58	U	SMPL
P4178-06MS - 1	3116MS	17155	3.8628	-	7470A	PPB	26 Sep 2024 11:36:14	U	SMPL
P4178-06MSD - 1	3116MSD	19461	4.3728	-	7470A	PPB	26 Sep 2024 11:38:31	U	SMPL
PB163649TB - 1	PB163649TB	-255	0.0125	-	7470A	PPB	26 Sep 2024 11:40:49	U	SMPL
PB163653TB - 1	PB163653TB	-35	0.0612	-	7470A	PPB	26 Sep 2024 11:43:06	U	SMPL
CCV23 - 1	CCV23	21868	4.9052	-	7470A	PPB	26 Sep 2024 11:45:21	U	SMPL
CCB23 - 1	CCB23	-231	0.0179	-	7470A	PPB	26 Sep 2024 11:47:37	U	SMPL
PB163673BL - 1	PBW	0	0.0689	-	7470A	PPB	26 Sep 2024 11:49:54	U	SMPL
PB163673BS - 1	LCSW	18618	4.1864	-	7470A	PPB	26 Sep 2024 11:52:10	U	SMPL
P4148-02LX5 - 1		-89	0.0493	-	7470A	PPB	26 Sep 2024 11:54:25	U	SMPL
P4148-02A - 1		18019	4.0539	-	7470A	PPB	26 Sep 2024 11:56:42	U	SMPL
P4178-06LX5 - 1		-50	0.0579	-	7470A	PPB	26 Sep 2024 11:58:58	U	SMPL
P4178-06A - 1		19489	4.3790	-	7470A	PPB	26 Sep 2024 12:01:15	U	SMPL
CCV24 - 1	CCV24	20109	4.5161	-	7470A	PPB	26 Sep 2024 12:03:32	U	SMPL
CCB24 - 1	CCB24	-172	0.0309	-	7470A	PPB	26 Sep 2024 12:05:51	U	SMPL