

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132608

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/26/2024	08:20
2	CAL2	1	Water	NA	NA	20.2	7.00	09/26/2024	08:21
3	CAL3	1	Water	NA	NA	20.3	10.02	09/26/2024	08:22
4	ICV	1	Water	NA	NA	20.3	6.98	09/26/2024	08:25
5	CCV1	1	Water	NA	NA	20.3	2.01	09/26/2024	08:30
6	P4191-03	1	Water	NA	NA	20.2	7.26	09/26/2024	08:35
7	P4194-01	1	Water	NA	NA	20.7	7.07	09/26/2024	08:40
8	P4194-01DUP	1	Water	NA	NA	20.8	7.08	09/26/2024	08:41
9	CCV2	1	Water	NA	NA	20.2	12.02	09/26/2024	08:45

WORKLIST(Hardcopy Internal Chain)

LB 132608

WorkList Name : ph p4199 water WorkList ID : 183833 Department : Wet-Chemistry Date : 09-26-2024 08:07:51

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4191-03	D HOPPER-1	Water	pH	Cool 4 deg C	PSEG03	J51	09/25/2024	9040C
P4194-01	T MANHOLE-119-KEARNY	Water	pH	Cool 4 deg C	PSEG03	J51	09/25/2024	9040C

Date/Time 09/26/24 08:10
 Raw Sample Received by: JH WOC
 Raw Sample Relinquished by: CP SM

Date/Time 09/26/24 11:00
 Raw Sample Received by: CP SM
 Raw Sample Relinquished by: JH WOC