

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132609

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/26/2024	08:20
2	CAL2	1	Water	NA	NA	20.2	7.00	09/26/2024	08:21
3	CAL3	1	Water	NA	NA	20.3	10.02	09/26/2024	08:22
4	ICV	1	Water	NA	NA	20.3	6.98	09/26/2024	08:25
5	CCV1	1	Water	NA	NA	20.2	2.01	09/26/2024	08:50
6	P4190-04	1	Solid	20.02	20	22.1	5.94	09/26/2024	09:00
7	P4190-04DUP	1	Solid	20.03	20	22.2	5.95	09/26/2024	09:01
8	P4192-02	1	Solid	20.04	20	22.7	7.96	09/26/2024	09:10
9	P4192-04	1	Solid	20.02	20	22.4	8.20	09/26/2024	09:15
10	CCV2	1	Water	NA	NA	20.3	12.02	09/26/2024	09:20

WORKLIST(Hardcopy Internal Chain)

VB132609

WorkList Name : corrosivity p4190 WorkList ID : 183834 Department : Wet-Chemistry Date : 09-26-2024 08:10:18

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4190-04	TP-T	Solid	Corrosivity	Cool 4 deg C	PSEG03	J41	09/25/2024	9045D
P4192-02	VNJ-222	Solid	Corrosivity	Cool 4 deg C	PSEG03	J51	09/25/2024	9045D
P4192-04	ETGI-355	Solid	Corrosivity	Cool 4 deg C	PSEG03	J51	09/25/2024	9045D

Date/Time 09/26/24 08:16
Raw Sample Received by: JH WDC
Raw Sample Relinquished by: CP SM

Date/Time 09/26/24 11:30
Raw Sample Received by: CP SM
Raw Sample Relinquished by: JH WDC