

LB132621


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43597



A= 0.0000e+000

B= 2.3362e-004 *slope*C= -1.6792e-001 *y-intercept*

Rho= 0.9999714

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	-0.035	-0.035	567	0.000	567					
0.05	0.050					442	0				
0.20	0.200	0.249	0.049	1784	0.0 %	1784					
2.50	2.500	2.504	0.004	11436	0.0 %	11436					
5.00	5.000	4.975	-0.025	22015	0.0 %	22015					
7.50	7.500	7.490	-0.010	32780	0.0 %	32780					
10.0	10.000	10.017	0.017	43597	0.0 %	43597					

25
0
0
0
0
0

LB132621

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	567	-		0 SFAM01.1	PPB	9/26/2024 14:10	Std
	0.2 S01	1784	-		0.2 SFAM01.1	PPB	9/26/2024 14:12	Std
	2.5 S02	11436	-		2.5 SFAM01.1	PPB	9/26/2024 14:14	Std
	5 S03	22015	-		5 SFAM01.1	PPB	9/26/2024 14:17	Std
	7.5 S04	32780	-		7.5 SFAM01.1	PPB	9/26/2024 14:19	Std
	10 S05	43597	-		10 SFAM01.1	PPB	9/26/2024 14:29	Std
ICV008	ICV008	17407	3.8987 -		SFAM01.1	PPB	9/26/2024 14:35	SMPL
ICB008	ICB008	542	-0.0413 -		SFAM01.1	PPB	9/26/2024 14:37	SMPL
CCV024	CCV024	21287	4.8052 -		SFAM01.1	PPB	9/26/2024 14:39	SMPL
CCB024	CCB024	547	-0.0401 -		SFAM01.1	PPB	9/26/2024 14:42	SMPL
PB163713BL	PBW713	458	-0.0609 -		SFAM01.1	PPB	9/26/2024 14:44	SMPL
P3945-01	MCOAN7	1012	0.0685 -		SFAM01.1	PPB	9/26/2024 14:46	SMPL
P3945-02	MCOAN7D	1028	0.0722 -		SFAM01.1	PPB	9/26/2024 14:48	SMPL
P3945-03	MCOAN7S	5386	1.0904 -		SFAM01.1	PPB	9/26/2024 14:53	SMPL
P3945-05	MCOAT3	4861	0.9677 -		SFAM01.1	PPB	9/26/2024 14:56	SMPL
P3945-06	MCOAT5	2046	0.3101 -		SFAM01.1	PPB	9/26/2024 14:58	SMPL
P3945-07	MCOAW5	449	-0.063 -		SFAM01.1	PPB	9/26/2024 15:00	SMPL
P3945-08	MCOAQ9	726	0.0017 -		SFAM01.1	PPB	9/26/2024 15:02	SMPL
P3945-09	MCOAR3	1972	0.2928 -		SFAM01.1	PPB	9/26/2024 15:05	SMPL
P3945-12	MCOAS5	904	0.0433 -		SFAM01.1	PPB	9/26/2024 15:07	SMPL
P3945-14	MCOAN6	695	-0.0055 -		SFAM01.1	PPB	9/26/2024 15:09	SMPL
P3945-15	MCOAN8	561	-0.0369 -		SFAM01.1	PPB	9/26/2024 15:12	SMPL
P3945-16	MCOAN9	769	0.0117 -		SFAM01.1	PPB	9/26/2024 15:14	SMPL
P3945-17	MCOAPO	629	-0.021 -		SFAM01.1	PPB	9/26/2024 15:16	SMPL
P3945-18	MCOAP1	542	-0.0413 -		SFAM01.1	PPB	9/26/2024 15:18	SMPL
P3945-19	MCOAQ3	1486	0.1792 -		SFAM01.1	PPB	9/26/2024 15:21	SMPL
P3945-20	MCOAQ7	491	-0.0532 -		SFAM01.1	PPB	9/26/2024 15:23	SMPL
P3945-21	MCOAR7	1927	0.2823 -		SFAM01.1	PPB	9/26/2024 15:25	SMPL
P3945-22	MCOAS1	377	-0.0798 -		SFAM01.1	PPB	9/26/2024 15:27	SMPL
P3945-23	MCOAS3	1776	0.247 -		SFAM01.1	PPB	9/26/2024 15:30	SMPL
CCV025	CCV025	20577	4.6393 -		SFAM01.1	PPB	9/26/2024 15:32	SMPL
CCB025	CCB025	650	-0.0161 -		SFAM01.1	PPB	9/26/2024 15:34	SMPL
P3945-24	MCOAP7	876	0.0367 -		SFAM01.1	PPB	9/26/2024 15:37	SMPL
P3945-25	MCOAP9	976	0.0601 -		SFAM01.1	PPB	9/26/2024 15:39	SMPL
P3945-26	MCOAQ1	1102	0.0895 -		SFAM01.1	PPB	9/26/2024 15:41	SMPL
PB163714BL	PBW714	950	0.054 -		SFAM01.1	PPB	9/26/2024 15:43	SMPL
P3946-01	MCOAT4	1455	0.172 -		SFAM01.1	PPB	9/26/2024 15:46	SMPL
P3946-02	MCOAT6	697	-0.0051 -		SFAM01.1	PPB	9/26/2024 15:48	SMPL
P3946-03	MCOAR4	1291	0.1337 -		SFAM01.1	PPB	9/26/2024 15:50	SMPL
P3946-04	MCOAR6	1700	0.2292 -		SFAM01.1	PPB	9/26/2024 15:52	SMPL
P3946-05	MCOAS0	784	0.0152 -		SFAM01.1	PPB	9/26/2024 15:55	SMPL
P3946-06	MCOAS6	678	-0.0095 -		SFAM01.1	PPB	9/26/2024 15:57	SMPL
P3946-07	MCOAS8	449	-0.063 -		SFAM01.1	PPB	9/26/2024 15:59	SMPL
P3946-08	MCOAQ8	451	-0.0626 -		SFAM01.1	PPB	9/26/2024 16:02	SMPL
P3946-09	MCOAR8	1465	0.1743 -		SFAM01.1	PPB	9/26/2024 16:04	SMPL
P3946-10	MCOAS2	424	-0.0689 -		SFAM01.1	PPB	9/26/2024 16:06	SMPL

LB132621

INSTRUMENT ID : CV1

P3946-11	MCOAS4	1412	0.162 -	SFAM01.1	PPB	9/26/2024 16:08 SMPL
P3946-12	MCOAQ0	795	0.0178 -	SFAM01.1	PPB	9/26/2024 16:11 SMPL
P3946-13	MCOAQ2	492	-0.053 -	SFAM01.1	PPB	9/26/2024 16:13 SMPL
P3946-14	MCOAQ4	795	0.0178 -	SFAM01.1	PPB	9/26/2024 16:15 SMPL
P3946-15	MCOAT0	511	-0.0485 -	SFAM01.1	PPB	9/26/2024 16:17 SMPL
P3946-16	MCOAP8	455	-0.0616 -	SFAM01.1	PPB	9/26/2024 16:20 SMPL
CCV026	CCV026	20614	4.648 -	SFAM01.1	PPB	9/26/2024 16:22 SMPL
CCB026	CCB026	390	-0.0768 -	SFAM01.1	PPB	9/26/2024 16:24 SMPL
P3946-17	MCOAQ6	845	0.0295 -	SFAM01.1	PPB	9/26/2024 16:26 SMPL
P3946-18	MCOAR1	602	-0.0273 -	SFAM01.1	PPB	9/26/2024 16:29 SMPL
P3946-19	MCOAR1D	747	0.0066 -	SFAM01.1	PPB	9/26/2024 16:31 SMPL
P3946-20	MCOAR1S	4258	0.8268 -	SFAM01.1	PPB	9/26/2024 16:35 SMPL
CCV027	CCV027	20256	4.5643 -	SFAM01.1	PPB	9/26/2024 16:38 SMPL
CCB207	CCB027	338	-0.089 -	SFAM01.1	PPB	9/26/2024 16:40 SMPL