

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB132623

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	09/26/2024	14:30
2	CAL2	1	Water	NA	NA	20.3	7.01	09/26/2024	14:31
3	CAL3	1	Water	NA	NA	20.3	10.02	09/26/2024	14:33
4	ICV	1	Water	NA	NA	20.3	7.02	09/26/2024	14:36
5	CCV1	1	Water	NA	NA	20.2	2.01	09/26/2024	14:40
6	P4202-01	1	Water	NA	NA	20.5	5.60	09/26/2024	14:45
7	P4202-01DUP	1	Water	NA	NA	20.6	5.61	09/26/2024	14:46
8	P4204-02	1	Water	NA	NA	20.1	2.89	09/26/2024	14:53
9	P4209-01	1	Water	NA	NA	20.3	6.42	09/26/2024	15:00
10	P4209-02	1	Water	NA	NA	20.4	6.46	09/26/2024	15:05
11	CCV2	1	Water	NA	NA	20.3	12.02	09/26/2024	15:10

WORKLIST(Hardcopy Internal Chain)

W132623

WorkList Name : ph p4209 WorkList ID : 183876 Department : Wet-Chemistry Date : 09-26-2024 14:18:59

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4202-01	C 50415	Water	pH	Cool 4 deg C	PSEG03	J41	09/26/2024	9040C
P4204-02	G BP-RW5FRACTANKSED-2024C	Water	pH	Cool 4 deg C	TETR06	J31	09/25/2024	9040C
P4209-01	H DRUM-1	Water	pH	Cool 4 deg C	PSEG03	J41	09/26/2024	9040C
P4209-02	L DRUM-2	Water	pH	Cool 4 deg C	PSEG03	J41	09/26/2024	9040C

Date/Time 09/26/24 14:20
 Raw Sample Received by: JH CWC
 Raw Sample Relinquished by: JH SM

Date/Time 09/26/24 16:00
 Raw Sample Received by: JH SM
 Raw Sample Relinquished by: JH CWC