

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132629

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	09/27/2024	10:45
2	P4172-03		1	11	100.00	>212.0	>212.0	09/27/2024	11:15
3	P4191-01		1	12	100.00	>212.0	>212.0	09/27/2024	11:45
4	P4191-01DUP		1	12	100.00	>212.0	>212.0	09/27/2024	12:15

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132629

WorkList Name : FP S9-26 WorkList ID : 183871 Department : Wet-Chemistry Date : 09-26-2024 10:47:40

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4172-03	OIL-TOTE-0924	Solid	Flash Point	Cool 4 deg C	PSEG03	J41	09/24/2024	1010B
P4191-01	NWB-2100	Solid	Flash Point	Cool 4 deg C	PSEG03	J51	09/25/2024	1010B

Date/Time 09/27/2024 10:35
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC

Date/Time 09/27/2024 12:55
 Raw Sample Received by: RM CWC
 Raw Sample Relinquished by: RM CWC