

7471B


LB132635

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51482



A= 0.0000e+000

B= 1.8973e-004

C= 1.8422e-002

Rho= 0.9991400

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%RSD
0.0	0.000	0.027	0.027	47	0.000	47					
0.2	0.200	0.179	-0.021	848	0.0 %	848					
2.5	2.500	2.404	-0.096	12574	0.0 %	12574					
5.0	5.000	5.013	0.013	26322	0.0 %	26322					
7.5	7.500	7.790	0.290	40962	0.0 %	40962					
10.0	10.000	9.786	-0.214	51482	0.0 %	51482					

111
140
12

LB132635 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 27 Sep 2024 09:07:13

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	SD	47	-	0.0000	7471B	PPB	27 Sep 2024 09:40:44	S	Std
0.2 - 1	SD	848	-	0.2000	7471B	PPB	27 Sep 2024 09:43:01	S	Std
2.5 - 1	SD	12574	-	2.5000	7471B	PPB	27 Sep 2024 09:45:19	S	Std
5.0 - 1	SD	26322	-	5.0000	7471B	PPB	27 Sep 2024 09:47:36	S	Std
7.5 - 1	SD	40962	-	7.5000	7471B	PPB	27 Sep 2024 09:49:54	S	Std
10.0 - 1	SD	55005	-	10.0000	7471B	PPB	27 Sep 2024 09:52:12	S	Std
10.0 - 1	SD	51482	-	10.0000	7471B	PPB	27 Sep 2024 09:56:05	S	Std
ICV10 - 1	ICV10	20670	3.9402	-7471B	PPB	PPB	27 Sep 2024 09:58:47	U	SMPL
ICB10 - 1	ICB10	-214	-0.0222	-7471B	PPB	PPB	27 Sep 2024 10:01:03	U	SMPL
CCV28 - 1	CCV28	27594	5.2539	-7471B	PPB	PPB	27 Sep 2024 10:03:22	U	SMPL
CCB28 - 1	CCB28	-238	-0.0267	-7471B	PPB	PPB	27 Sep 2024 10:05:37	U	SMPL
CRA - 1	CRA	847	0.1791	-7471B	PPB	PPB	27 Sep 2024 10:07:55	U	SMPL
HighStd - 1	HighStd	53939	10.2525	-7471B	PPB	PPB	27 Sep 2024 10:10:10	U	SMPL
ChkStd - 1	ChkStd	37826	7.1953	-7471B	PPB	PPB	27 Sep 2024 10:12:25	U	SMPL
PB163731BL - 1	PBS	-204	-0.0203	-7471B	PPB	PPB	27 Sep 2024 10:14:46	U	SMPL
P4190-01 - 1	TP-T	6801	1.3088	-7471B	PPB	PPB	27 Sep 2024 10:19:22	U	SMPL
P4190-01DUP - 1	TP-TDUP	4813	0.9316	-7471B	PPB	PPB	27 Sep 2024 10:21:41	U	SMPL
P4190-01MS - 1	TP-TMS	29975	5.7057	-7471B	PPB	PPB	27 Sep 2024 10:24:00	U	SMPL
P4190-01MSD - 1	TP-TMSD	31593	6.0127	-7471B	PPB	PPB	27 Sep 2024 10:26:18	U	SMPL
P4191-01 - 1	NWB-2100	11621	2.2233	-7471B	PPB	PPB	27 Sep 2024 10:28:37	U	SMPL
CCV29 - 1	CCV29	26565	5.0587	-7471B	PPB	PPB	27 Sep 2024 10:30:59	U	SMPL
CCB29 - 1	CCB29	-112	-0.0028	-7471B	PPB	PPB	27 Sep 2024 10:33:17	U	SMPL
P4192-01 - 1	VNJ-222	3034	0.5941	-7471B	PPB	PPB	27 Sep 2024 10:35:35	U	SMPL
P4192-03 - 1	ETGI-355	3193	0.6242	-7471B	PPB	PPB	27 Sep 2024 10:43:31	U	SMPL
P4197-01 - 1	EO-01-092524	7406	1.4236	-7471B	PPB	PPB	27 Sep 2024 10:45:46	U	SMPL
PB163731BS - 1	LCSS	19547	3.7272	-7471B	PPB	PPB	27 Sep 2024 10:48:22	U	SMPL
P4190-01LX5 - 1		1533	0.3093	-7471B	PPB	PPB	27 Sep 2024 10:51:08	U	SMPL
P4190-01A - 1		25215	4.8026	-7471B	PPB	PPB	27 Sep 2024 10:53:23	U	SMPL
CCV30 - 1	CCV30	26759	5.0955	-7471B	PPB	PPB	27 Sep 2024 10:55:39	U	SMPL
CCB30 - 1	CCB30	-114	-0.0032	-7471B	PPB	PPB	27 Sep 2024 10:57:58	U	SMPL