

LB132645


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

42775



A= 0.0000e+000

B= 2.3716e-004 slope

C= -1.7871e-001 y-intercept

Rho= 0.9999264

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	o/n
0.00	0.000	-0.043	-0.043	571	0.000	571					
0.05	0.050					442	0				
0.20	0.200	0.252	0.052	1815	0.0 %	1815					
2.50	2.500	2.494	-0.006	11268	0.0 %	11268					
5.00	5.000	4.962	-0.038	21677	0.0 %	21677					
7.50	7.500	7.570	0.070	32671	0.0 %	32671					
10.0	10.000	9.966	-0.034	42775	0.0 %	42775					

0/00
= 26
0
- 1
1
0

LB132645

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	571	-		0 SFAM01.1	PPB	9/27/2024 13:22	Std
	0.2 S01	1815	-		0.2 SFAM01.1	PPB	9/27/2024 13:24	Std
	2.5 S02	11268	-		2.5 SFAM01.1	PPB	9/27/2024 13:27	Std
	5 S03	21677	-		5 SFAM01.1	PPB	9/27/2024 13:29	Std
	7.5 S04	32671	-		7.5 SFAM01.1	PPB	9/27/2024 13:31	Std
	10 S05	42775	-		10 SFAM01.1	PPB	9/27/2024 13:36	Std
ICV011	ICV011	18205	4.1389 -		SFAM01.1	PPB	9/27/2024 13:45	SMPL
ICB011	ICB011	508	-0.0582 -		SFAM01.1	PPB	9/27/2024 13:50	SMPL
CCV031	CCV031	22013	5.042 -		SFAM01.1	PPB	9/27/2024 13:52	SMPL
CCB031	CCB031	369	-0.0912 -		SFAM01.1	PPB	9/27/2024 13:54	SMPL
PB163753BL	PBW753	569	-0.0438 -		SFAM01.1	PPB	9/27/2024 14:00	SMPL
P3943-01	MH0BC6	104	-0.154 -		SFAM01.1	PPB	9/27/2024 14:02	SMPL
P3943-02	MH0BC8	21	-0.1737 -		SFAM01.1	PPB	9/27/2024 14:05	SMPL
P3943-03	MH0BD0	121	-0.15 -		SFAM01.1	PPB	9/27/2024 14:07	SMPL
P3943-04	MH0BD2	41	-0.169 -		SFAM01.1	PPB	9/27/2024 14:09	SMPL
P3943-05	MH0BD2D	10	-0.1763 -		SFAM01.1	PPB	9/27/2024 14:11	SMPL
P3943-06	MH0BD2S	5404	1.1029 -		SFAM01.1	PPB	9/27/2024 14:14	SMPL
P3943-07	MH0BD4	-34	-0.1868 -		SFAM01.1	PPB	9/27/2024 14:16	SMPL
P3943-08	MH0BD6	27	-0.1723 -		SFAM01.1	PPB	9/27/2024 14:18	SMPL
P3943-09	MH0BD8	31	-0.1714 -		SFAM01.1	PPB	9/27/2024 14:21	SMPL
P3943-10	MH0BE0	41	-0.169 -		SFAM01.1	PPB	9/27/2024 14:23	SMPL
P3943-11	MH0BE2	18	-0.1744 -		SFAM01.1	PPB	9/27/2024 14:25	SMPL
P3943-12	MH0BE4	29	-0.1718 -		SFAM01.1	PPB	9/27/2024 14:27	SMPL
P3943-13	MH0BE6	69	-0.1623 -		SFAM01.1	PPB	9/27/2024 14:30	SMPL
PB163754BL	PBW754	106	-0.1536 -		SFAM01.1	PPB	9/27/2024 14:32	SMPL
P3944-01	MH0BC7	312	-0.1047 -		SFAM01.1	PPB	9/27/2024 14:34	SMPL
P3944-02	MH0BC9	43	-0.1685 -		SFAM01.1	PPB	9/27/2024 14:36	SMPL
P3944-03	MH0BD1	59	-0.1647 -		SFAM01.1	PPB	9/27/2024 14:39	SMPL
P3944-04	MH0BD3	85	-0.1586 -		SFAM01.1	PPB	9/27/2024 14:41	SMPL
P3944-05	MH0BD3D	112	-0.1522 -		SFAM01.1	PPB	9/27/2024 14:43	SMPL
CCV032	CCV032	21473	4.9139 -		SFAM01.1	PPB	9/27/2024 14:45	SMPL
CCB032	CCB032	377	-0.0893 -		SFAM01.1	PPB	9/27/2024 14:48	SMPL
P3944-06	MH0BD3S	5832	1.2044 -		SFAM01.1	PPB	9/27/2024 14:50	SMPL
P3944-07	MH0BD5	62	-0.164 -		SFAM01.1	PPB	9/27/2024 14:52	SMPL
P3944-08	MH0BD7	152	-0.1427 -		SFAM01.1	PPB	9/27/2024 14:54	SMPL
P3944-09	MH0BD9	28	-0.1721 -		SFAM01.1	PPB	9/27/2024 14:57	SMPL
P3944-10	MH0BE1	75	-0.1609 -		SFAM01.1	PPB	9/27/2024 14:59	SMPL
P3944-11	MH0BE3	89	-0.1576 -		SFAM01.1	PPB	9/27/2024 15:01	SMPL
P3944-12	MH0BE5	43	-0.1685 -		SFAM01.1	PPB	9/27/2024 15:03	SMPL
P3944-13	MH0BE7	34	-0.1707 -		SFAM01.1	PPB	9/27/2024 15:06	SMPL
PB163755BL	PB755	217	-0.1272 -		SFAM01.1	PPB	9/27/2024 15:12	SMPL
P4091-01	MCOB15	49	-0.1671 -		SFAM01.1	PPB	9/27/2024 15:15	SMPL
P4091-02	MCOB15D	86	-0.1583 -		SFAM01.1	PPB	9/27/2024 15:17	SMPL
P4091-04	MCOB16	66	-0.1631 -		SFAM01.1	PPB	9/27/2024 15:21	SMPL
P4091-05	MCOB17	2228	0.3497 -		SFAM01.1	PPB	9/27/2024 15:24	SMPL
P4091-06	MCOB18	102	-0.1545 -		SFAM01.1	PPB	9/27/2024 15:26	SMPL

LB132645

INSTRUMENT ID : CV1

P4091-07	MC0B19	115	-0.1514 -	SFAM01.1	PPB	9/27/2024 15:28 SMPL
P4091-08	MC0B28	21	-0.1737 -	SFAM01.1	PPB	9/27/2024 15:30 SMPL
P4091-09	MC0B36	54	-0.1659 -	SFAM01.1	PPB	9/27/2024 15:33 SMPL
P4091-10	MC0B37	40	-0.1692 -	SFAM01.1	PPB	9/27/2024 15:35 SMPL
P4091-11	MC0B47	41	-0.169 -	SFAM01.1	PPB	9/27/2024 15:37 SMPL
CCV033	CCV033	21137	4.8342 -	SFAM01.1	PPB	9/27/2024 15:40 SMPL
CCB033	CCB033	325	-0.1016 -	SFAM01.1	PPB	9/27/2024 15:42 SMPL
P4091-12	MC0B53	258	-0.1175 -	SFAM01.1	PPB	9/27/2024 15:44 SMPL
P4091-13	MC0B54	-3	-0.1794 -	SFAM01.1	PPB	9/27/2024 15:46 SMPL
P4091-03	MC0B15S	5291	1.0761 -	SFAM01.1	PPB	9/27/2024 15:49 SMPL
CCV034	CCV034	21475	4.9144 -	SFAM01.1	PPB	9/27/2024 15:51 SMPL
CCB034	CCB034	447	-0.0727 -	SFAM01.1	PPB	9/27/2024 16:02 SMPL