

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132652

Ambient Barometric Pressure (mmHg): 765.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3087

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	82.1	09/27/2024	15:05
2	P4191-03		1	11	100.00	>212.0	>212.0	09/27/2024	15:35
3	P4191-03DUP		1	11	100.00	>212.0	>212.0	09/27/2024	16:05
4	P4194-01		1	14	100.00	>212.0	>212.0	09/27/2024	16:35
5	P4202-01		1	12	100.00	>212.0	>212.0	09/27/2024	17:05
6	P4204-02		1	11	100.00	>212.0	>212.0	09/27/2024	17:35

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

16132652

WorkList Name : FP-9-26 WorkList ID : 183872 Department : Wet-Chemistry Date : 09-26-2024 12:49:55

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4191-03 F	HOPPER-1	Water	Flash Point	Cool 4 deg C	PSEG03	J51	09/25/2024	1010B
P4194-01 N	MANHOLE-119-KEARNY	Water	Flash Point	Cool 4 deg C	PSEG03	J51	09/25/2024	1010B
P4202-01 E	50415	Water	Flash Point	Cool 4 deg C	PSEG03	J41	09/26/2024	1010B
P4204-02 G	BP-RW5FRACTANKSED-2024C	Water	Flash Point	Cool 4 deg C	TETR06	J31	09/25/2024	1010B

Date/Time 09/27/2024 14:55
 Raw Sample Received by: RM CWS
 Raw Sample Relinquished by: M. Hall

Date/Time 09/27/2024 18:10
 Raw Sample Received by: JPL CWS
 Raw Sample Relinquished by: RM CWS