

Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB132660

Reagent/Standard	Lot/Log #
Turbidity Calibration std, 0NTU	WP109925
Turbidity Calibration std, 20NTU	WP109927
Turbidity Calibration std, 40NTU	WP109926
Turbidity Calibration std, 1NTU	WP109929
Turbidity Calibration std, 5NTU	WP109928
Turbidity Calibration - CCV std, 10 NTU	WP109930
10 NTU Standard 500 ml	W3116

Intercept: -0.0156

Slope: 0.9363

Regression: 0.9999

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (mg/l)	%D	Anal Date	Anal Time
1	CAL1	0	1	0.148	0.18		09/26/2024	10:24
2	CAL2	1	1	1.020	1.11	10.6	09/26/2024	10:28
3	CAL3	5	1	4.510	4.83	-3.3	09/26/2024	10:32
4	CAL4	10	1	9.390	10.05	0.4	09/26/2024	10:36
5	CAL5	20	1	18.400	19.67	-1.7	09/26/2024	10:40
6	CAL6	40	1	37.600	40.18	0.4	09/26/2024	10:44



Analytical Summary Report

Analysis Method: 180.1

ANALYST: Niha

Parameter: Turbidity

SUPERVISOR REVIEW BY: Iwona

Run Number: LB132660

Seq	Lab ID	True Value (NTU)	Dilution	Reading	Result (NTU)	AnalDate	Anal Time
1	ICV	10	1	8.910	9.533	09/26/2024	10:48
2	ICB		1	0.146	0.173	09/26/2024	10:52
3	CCV1	10	1	8.820	9.437	09/26/2024	10:56
4	CCB1		1	0.141	0.167	09/26/2024	11:00
5	LB132660BL		1	0.138	0.164	09/26/2024	11:04
6	P3848-34		1	24.100	25.756	09/26/2024	11:08
7	CCV2	10	1	9.300	9.949	09/26/2024	11:12
8	CCB2		1	0.140	0.166	09/26/2024	11:16

LB13 2660

WORKLIST(Hardcopy Internal Chain)

WorkList Name : TURBIDITY - 09262024

WorkList ID : 183878

Department : Wet-Chemistry

Date : 09-26-2024 09:04:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P3848-34	PT-TURB-WP	Water	Turbidity	Cool 4 deg C	CHEM02	QA Of	09/03/2024	180.1

Date/Time09. 26.2024. 09:15

Raw Sample Received by: NFWC

Raw Sample Relinquished by: SJ (QAO)

Date/TimeNA

Raw Sample Received by: y

Raw Sample Relinquished by: y