

LB132665


7471B

INSTRUMENT ID: CV1

Linear

 μ Abs.:

51330



A= 0.0000e+000

B= 1.9385e-004

C= -1.2452e-002

Rho= 0.9997813

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.0	0.000	0.008	0.008	107	0.000	107					4.5
0.2	0.200	0.227	0.027	1235	0.0 %	1235					14
2.5	2.500	2.469	-0.031	12803	0.0 %	12803					1
5.0	5.000	4.910	-0.090	25392	0.0 %	25392					2
7.5	7.500	7.648	0.148	39519	0.0 %	39519					2
10.0	10.000	9.938	-0.062	51330	0.0 %	51330					1

LB132665 INSTRUMENT ID : CV1

Method: 7471B

Operator: Admin

Date of Analysis: 30 Sep 2024 13:30:01

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	107	-	0.0000	7471B	PPB	30 Sep 2024 13:32:15	S	Std
0.2 - 1	50-2	1235	-	0.2000	7471B	PPB	30 Sep 2024 13:34:32	S	Std
2.5 - 1	52-5	12803	-	2.5000	7471B	PPB	30 Sep 2024 13:36:49	S	Std
5.0 - 1	55	25392	-	5.0000	7471B	PPB	30 Sep 2024 13:39:05	S	Std
7.5 - 1	57-5	39519	-	7.5000	7471B	PPB	30 Sep 2024 13:41:24	S	Std
10.0 - 1	51V	51330	-	10.0000	7471B	PPB	30 Sep 2024 13:43:43	S	Std
ICV15 - 1	ICV15	19330	3.7346	-	7471B	PPB	30 Sep 2024 13:46:43	U	SMPL
ICB15 - 1	ICB15	-266	-0.0640	-	7471B	PPB	30 Sep 2024 13:48:57	U	SMPL
CCV46 - 1	CCV46	25150	4.8628	-	7471B	PPB	30 Sep 2024 13:51:15	U	SMPL
CCB46 - 1	CCB46	-257	-0.0623	-	7471B	PPB	30 Sep 2024 13:53:30	U	SMPL
CRA - 1	CRA	1091	0.1990	-	7471B	PPB	30 Sep 2024 13:55:49	U	SMPL
HighStd - 1	HighStd	51531	9.9766	-	7471B	PPB	30 Sep 2024 13:58:04	U	SMPL
ChkStd - 1	ChkStd	36437	7.0507	-	7471B	PPB	30 Sep 2024 14:05:45	U	SMPL
PB163800BL - 1	PBS	-194	-0.0601	-	7471B	PPB	30 Sep 2024 14:08:00	U	SMPL
PB163800BS - 1	LCSS	22813	4.4097	-	7471B	PPB	30 Sep 2024 14:10:22	U	SMPL
P4198-01 - 1	TP2-WC-1	3681	0.7011	-	7471B	PPB	30 Sep 2024 14:12:38	U	SMPL
P4198-06 - 1	TP1-WC-1	3728	0.7102	-	7471B	PPB	30 Sep 2024 14:14:57	U	SMPL
P4198-11 - 1	TP1-5-WC-1	996	0.1806	-	7471B	PPB	30 Sep 2024 14:17:16	U	SMPL
P4198-16 - 1	TP3-WC-1	2734	0.5175	-	7471B	PPB	30 Sep 2024 14:19:36	U	SMPL
P4198-21 - 1	TP4-WC-1	1018	0.1849	-	7471B	PPB	30 Sep 2024 14:21:51	U	SMPL
CCV47 - 1	CCV47	26004	5.0283	-	7471B	PPB	30 Sep 2024 14:24:06	U	SMPL
CCB47 - 1	CCB47	-198	-0.0508	-	7471B	PPB	30 Sep 2024 14:26:22	U	SMPL
P4198-26 - 1	TP5-WC-1	6042	1.1588	-	7471B	PPB	30 Sep 2024 14:28:41	U	SMPL
P4212-01 - 1	TEST-PIT-NORTH	1462	0.2709	-	7471B	PPB	30 Sep 2024 14:30:57	U	SMPL
P4212-03 - 1	TEST-PIT-SOUTH	1482	0.2748	-	7471B	PPB	30 Sep 2024 14:33:18	U	SMPL
P4213-01 - 1	HD-02-092624	1299	0.2394	-	7471B	PPB	30 Sep 2024 14:35:35	U	SMPL
P4214-01 - 1	CONCRETE-WC-1	1241	0.2281	-	7471B	PPB	30 Sep 2024 14:37:51	U	SMPL
P4222-01 - 1	TP-5	2505	0.4731	-	7471B	PPB	30 Sep 2024 14:40:08	U	SMPL
P4225-01 - 1	TR-04-092724	1459	0.2704	-	7471B	PPB	30 Sep 2024 14:42:25	U	SMPL
P4225-01DUP - 1	TR-04-092724DUP	1764	0.3295	-	7471B	PPB	30 Sep 2024 14:44:41	U	SMPL
P4225-01MSD - 1	TR-04-092724MSD	20510	3.9633	-	7471B	PPB	30 Sep 2024 14:49:14	U	SMPL
CCV48 - 1	CCV48	26171	5.0607	-	7471B	PPB	30 Sep 2024 14:51:33	U	SMPL
CCB48 - 1	CCB48	-190	-0.0493	-	7471B	PPB	30 Sep 2024 14:53:52	U	SMPL
P4225-01LX5 - 1		259	0.0378	-	7471B	PPB	30 Sep 2024 14:56:11	U	SMPL
P4225-01A - 1		22839	4.4148	-	7471B	PPB	30 Sep 2024 14:58:27	U	SMPL
P4225-01MS - 1	TR-04-092724MS	21485	4.1523	-	7471B	PPB	30 Sep 2024 15:08:36	U	SMPL
CCV49 - 1	CCV49	26669	5.1572	-	7471B	PPB	30 Sep 2024 15:11:13	U	SMPL
CCB49 - 1	CCB49	-182	-0.0477	-	7471B	PPB	30 Sep 2024 15:13:28	U	SMPL