

L3132669


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

46231



A= 0.0000e+000

B= 2.2235e-004 slope

C= -2.4383e-001 y-intercept

Rho= 0.9995208

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	σ/μ
0.00	0.000	-0.006	-0.006	1071	0.000	1071					
0.05	0.050					442					
0.20	0.200	0.212	0.012	2048	0.0 %	2048	0				
2.50	2.500	2.423	-0.077	11992	0.0 %	11992					
5.00	5.000	5.205	0.205	24506	0.0 %	24506					
7.50	7.500	7.330	-0.170	34064	0.0 %	34064					
10.0	10.000	10.036	0.036	46231	0.0 %	46231					

σ/μ
= 6
-3
4
-2
0

LB132669
INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	1071	-		0 SFAM01.1	PPB	9/30/2024 15:54	Std
	0.2 S01	2048	-		0.2 SFAM01.1	PPB	9/30/2024 15:56	Std
	2.5 S02	11992	-		2.5 SFAM01.1	PPB	9/30/2024 15:59	Std
	5 S03	24506	-		5 SFAM01.1	PPB	9/30/2024 16:01	Std
	7.5 S04	34064	-		7.5 SFAM01.1	PPB	9/30/2024 16:03	Std
	10 S05	46231	-		10 SFAM01.1	PPB	9/30/2024 16:06	Std
ICV016	ICV016	18082	3.7768 -		SFAM01.1	PPB	9/30/2024 16:09	SMPL
ICB016	ICB016	973	-0.0275 -		SFAM01.1	PPB	9/30/2024 16:11	SMPL
CCV050	CCV050	23998	5.0922 -		SFAM01.1	PPB	9/30/2024 16:13	SMPL
CCB050	CCB050	736	-0.0802 -		SFAM01.1	PPB	9/30/2024 16:15	SMPL
PB163801BL	PBS801	796	-0.0668 -		SFAM01.1	PPB	9/30/2024 16:18	SMPL
P3939-01	ME20W0	1091	-0.0012 -		SFAM01.1	PPB	9/30/2024 16:20	SMPL
P3939-02	ME20W1	1110	0.003 -		SFAM01.1	PPB	9/30/2024 16:22	SMPL
P3939-03	ME20W2	973	-0.0275 -		SFAM01.1	PPB	9/30/2024 16:24	SMPL
P3939-04	ME20W5	849	-0.0551 -		SFAM01.1	PPB	9/30/2024 16:27	SMPL
P3939-05	ME20W6	1167	0.0157 -		SFAM01.1	PPB	9/30/2024 16:29	SMPL
P3939-06	ME20W7	1015	-0.0181 -		SFAM01.1	PPB	9/30/2024 16:31	SMPL
P3939-07	ME20W8	1119	0.005 -		SFAM01.1	PPB	9/30/2024 16:34	SMPL
P3939-08	ME20W9	1133	0.0081 -		SFAM01.1	PPB	9/30/2024 16:36	SMPL
P3939-09	ME20X0	964	-0.0295 -		SFAM01.1	PPB	9/30/2024 16:38	SMPL
P3939-10	ME20Y1	1492	0.0879 -		SFAM01.1	PPB	9/30/2024 16:40	SMPL
P3939-11	ME20Y2	1045	-0.0115 -		SFAM01.1	PPB	9/30/2024 16:43	SMPL
P3939-12	ME20Y3	1054	-0.0095 -		SFAM01.1	PPB	9/30/2024 16:45	SMPL
P3939-13	ME20Y4	1272	0.039 -		SFAM01.1	PPB	9/30/2024 16:47	SMPL
P3939-14	ME20Y5	1055	-0.0092 -		SFAM01.1	PPB	9/30/2024 16:49	SMPL
P3939-15	ME20Y6	1049	-0.0106 -		SFAM01.1	PPB	9/30/2024 16:52	SMPL
P3939-16	ME20Y7	1236	0.031 -		SFAM01.1	PPB	9/30/2024 16:54	SMPL
P3939-17	ME20Z9	926	-0.0379 -		SFAM01.1	PPB	9/30/2024 16:56	SMPL
P3939-18	ME20Z9D	1246	0.0332 -		SFAM01.1	PPB	9/30/2024 16:58	SMPL
P3939-19	ME20Z9S	10972	2.1958 -		SFAM01.1	PPB	9/30/2024 17:01	SMPL
CCV051	CCV051	24299	5.1592 -		SFAM01.1	PPB	9/30/2024 17:03	SMPL
CCB051	CCB051	993	-0.023 -		SFAM01.1	PPB	9/30/2024 17:05	SMPL
P3939-20	ME2100	487	-0.1355 -		SFAM01.1	PPB	9/30/2024 17:08	SMPL
P3939-21	ME2101	415	-0.1516 -		SFAM01.1	PPB	9/30/2024 17:10	SMPL
P3939-22	ME2102	423	-0.1498 -		SFAM01.1	PPB	9/30/2024 17:12	SMPL
PB163802BL	PBS802	1328	0.0515 -		SFAM01.1	PPB	9/30/2024 17:17	SMPL
P4006-01	MH2A01	1521	0.0944 -		SFAM01.1	PPB	9/30/2024 17:19	SMPL
P4006-02	MH2A02	1646	0.1222 -		SFAM01.1	PPB	9/30/2024 17:22	SMPL
P4006-03	MH2A03	1154	0.0128 -		SFAM01.1	PPB	9/30/2024 17:24	SMPL
P4006-04	MH2A04	1928	0.1849 -		SFAM01.1	PPB	9/30/2024 17:26	SMPL
P4006-05	MH2A05	1927	0.1846 -		SFAM01.1	PPB	9/30/2024 17:28	SMPL
P4006-06	MH2A06	1666	0.1266 -		SFAM01.1	PPB	9/30/2024 17:31	SMPL
P4006-07	MH2A07	5619	1.0056 -		SFAM01.1	PPB	9/30/2024 17:33	SMPL
P4006-08	MH2A08	1222	0.0279 -		SFAM01.1	PPB	9/30/2024 17:35	SMPL
P4006-09	MH2A09	114172	25.1428 -		SFAM01.1	PPB	9/30/2024 17:37	SMPL
P4006-10	MH2A10	113799	25.0599 -		SFAM01.1	PPB	9/30/2024 17:40	SMPL

LB132669

INSTRUMENT ID : CV1

P4006-11	MH2A11	125342	27.6265 -	SFAM01.1	PPB	9/30/2024 17:42 SMPL
P4006-12	MH2A12	73993	16.2088 -	SFAM01.1	PPB	9/30/2024 17:44 SMPL
P4006-13	MH2A13	105589	23.2344 -	SFAM01.1	PPB	9/30/2024 17:47 SMPL
P4006-14	MH2A14	573	-0.1164 -	SFAM01.1	PPB	9/30/2024 17:49 SMPL
P4006-15	MH2A15	70877	15.516 -	SFAM01.1	PPB	9/30/2024 17:51 SMPL
P4006-16	MH2A16	119444	26.3151 -	SFAM01.1	PPB	9/30/2024 17:54 SMPL
CCV052	CCV052	23907	5.072 -	SFAM01.1	PPB	9/30/2024 17:56 SMPL
CCB052	CCB052	908	-0.0419 -	SFAM01.1	PPB	9/30/2024 17:58 SMPL
P4006-17	MH2A17	10182	2.0202 -	SFAM01.1	PPB	9/30/2024 18:01 SMPL
P4006-18	MH2A18	14176	2.9083 -	SFAM01.1	PPB	9/30/2024 18:03 SMPL
P4006-19	MH2A18D	11262	2.2603 -	SFAM01.1	PPB	9/30/2024 18:05 SMPL
P4006-20	MH2A18S	22897	4.8474 -	SFAM01.1	PPB	9/30/2024 18:07 SMPL
P4006-21	MH2A19	11116	2.2279 -	SFAM01.1	PPB	9/30/2024 18:10 SMPL
CCV053	CCV053	24195	5.136 -	SFAM01.1	PPB	9/30/2024 18:23 SMPL
CCB053	CCB053	927	-0.0377 -	SFAM01.1	PPB	9/30/2024 18:26 SMPL