

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132670

Ambient Barometric Pressure (mmHg): 760.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	82.4	09/30/2024	10:25
2	P4176-01		1	12	100.00	>212.0	>212.0	09/30/2024	11:30
3	P4176-01DUP		1	12	100.00	>212.0	>212.0	09/30/2024	12:00
4	P4176-02		1	11	100.00	>212.0	>212.0	09/30/2024	13:45
5	P4176-03		1	14	100.00	>212.0	>212.0	09/30/2024	14:15
6	P4176-04		1	14	100.00	>212.0	>212.0	09/30/2024	14:45
7	P4176-05		1	11	100.00	>212.0	>212.0	09/30/2024	15:15
8	P4176-06		1	14	100.00	>212.0	>212.0	09/30/2024	15:45

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

6132670

WorkList Name : fp -9-30 WorkList ID : 183940 Department : Wet-Chemistry Date : 09-30-2024 08:15:54

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4176-01	FB-NC-SYS2	Water	Flash Point	Cool 4 deg C	RUTW01	J41	09/24/2024	1010B
P4176-02	FB-1C-SYS2	Water	Flash Point	Cool 4 deg C	RUTW01	J41	09/24/2024	1010B
P4176-03	FB-2C-SYS2	Water	Flash Point	Cool 4 deg C	RUTW01	J41	09/24/2024	1010B
P4176-04	FB-NC-SYS1	Water	Flash Point	Cool 4 deg C	RUTW01	J41	09/24/2024	1010B
P4176-05	FB-1C-SYS1	Water	Flash Point	Cool 4 deg C	RUTW01	J41	09/24/2024	1010B
P4176-06	FB-2C-SYS1	Water	Flash Point	Cool 4 deg C	RUTW01	J41	09/24/2024	1010B

Date/Time 09/30/2024 10:15
Raw Sample Received by: RM M. W. Lee
Raw Sample Relinquished by: M. W. Lee

Date/Time 09/30/2024 16:15
Raw Sample Received by: M. W. Lee
Raw Sample Relinquished by: RM W. Lee