

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132671

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	9	28.00	82.4	81.7	10/01/2024	10:00
2	P4209-01		1	12	100.00	>212.0	>212.0	10/01/2024	10:30
3	P4209-01DUP		1	12	100.00	>212.0	>212.0	10/01/2024	11:00
4	P4209-02		1	11	100.00	>212.0	>212.0	10/01/2024	11:30
5	P4235-02		1	14	100.00	>212.0	>212.0	10/01/2024	12:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

WORKLIST(Hardcopy Internal Chain)

66132671

WorkList Name : fp-10-1

WorkList ID : 183983

Department : Wet-Chemistry

Date : 10-01-2024 08:31:31

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4209-01	DRUM-1	Water	Flash Point	Cool 4 deg C	PSEG03	J41	09/26/2024	1010B
P4209-02	DRUM-2	Water	Flash Point	Cool 4 deg C	PSEG03	J41	09/26/2024	1010B
P4235-02	A1403	Water	Flash Point	Cool 4 deg C	PSEG03	J51	09/30/2024	1010B

Date/Time 10/01/2024 09:50
Raw Sample Received by: RM CWS
Raw Sample Relinquished by: RM CWS

Date/Time 10/01/2024 12:40
Raw Sample Received by: RM CWS
Raw Sample Relinquished by: RM CWS