

LB132675


SFAM01.1

INSTRUMENT ID: CV1

Linear

μ Abs.:

44774



A= 0.0000e+000
 B= 2.2914e-004 slope
 C= -1.9458e-001 y-intercept
 Rho= 0.9998842
 Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	
0.00	0.000	0.008	0.008	886	0.000	886					
0.05	0.050					442	0				
0.20	0.200	0.260	0.060	1985	0.0 %	1985					
2.50	2.500	2.429	-0.071	11449	0.0 %	11449					
5.00	5.000	5.014	0.014	22729	0.0 %	22729					
7.50	7.500	7.424	-0.076	33249	0.0 %	33249					
10.0	10.000	10.065	0.065	44774	0.0 %	44774					

0/0
 11/3/18
 1-0301

LB132675

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	886	-		0 SFAM01.1	PPB	10/1/2024 8:51	Std
	0.2 S01	1985	-		0.2 SFAM01.1	PPB	10/1/2024 8:53	Std
	2.5 S02	11449	-		2.5 SFAM01.1	PPB	10/1/2024 8:56	Std
	5 S03	22729	-		5 SFAM01.1	PPB	10/1/2024 8:58	Std
	7.5 S04	33249	-		7.5 SFAM01.1	PPB	10/1/2024 9:00	Std
	10 S05	44774	-		10 SFAM01.1	PPB	10/1/2024 9:02	Std
ICV17	ICV17	17884	3.9033 -		SFAM01.1	PPB	10/1/2024 9:05	SMPL
ICB17	ICB17	637	-0.0486 -		SFAM01.1	PPB	10/1/2024 9:08	SMPL
CCV54	CCV54	23412	5.17 -		SFAM01.1	PPB	10/1/2024 9:10	SMPL
CCB54	CCB54	636	-0.0488 -		SFAM01.1	PPB	10/1/2024 9:12	SMPL
PB163803BL	PB163803BL	906	0.013 -		SFAM01.1	PPB	10/1/2024 9:14	SMPL
P4008-01	MH2A20	43697	9.8181 -		SFAM01.1	PPB	10/1/2024 9:17	SMPL
P4008-02	MH2A20D	41787	9.3804 -		SFAM01.1	PPB	10/1/2024 9:19	SMPL
P4008-03	MH2A20S	54300	12.2477 -		SFAM01.1	PPB	10/1/2024 9:21	SMPL
P4008-04	MH2A21	57151	12.9009 -		SFAM01.1	PPB	10/1/2024 9:24	SMPL
P4006-09DLX5	MH2A09	25505	5.6496 -		SFAM01.1	PPB	10/1/2024 9:26	SMPL
P4006-10DLX5	MH2A10	26500	5.8776 -		SFAM01.1	PPB	10/1/2024 9:28	SMPL
P4006-11DLX5	MH2A11	25320	5.6072 -		SFAM01.1	PPB	10/1/2024 9:30	SMPL
P4006-12DLX5	MH2A12	16715	3.6355 -		SFAM01.1	PPB	10/1/2024 9:33	SMPL
P4006-13DLX5	MH2A13	21861	4.8146 -		SFAM01.1	PPB	10/1/2024 9:38	SMPL
P4006-15DLX5	MH2A15	13452	2.8878 -		SFAM01.1	PPB	10/1/2024 9:43	SMPL
P4006-16DLX5	MH2A16	27462	6.098 -		SFAM01.1	PPB	10/1/2024 9:45	SMPL
P4008-01DLX2	MH2A20	20767	4.5639 -		SFAM01.1	PPB	10/1/2024 9:48	SMPL
P4008-02DLX2	MH2A20D	20917	4.5983 -		SFAM01.1	PPB	10/1/2024 9:50	SMPL
P4008-03DLX2	MH2A20S	28220	6.2717 -		SFAM01.1	PPB	10/1/2024 9:52	SMPL
P4008-04DLX2	MH2A21	29488	6.5623 -		SFAM01.1	PPB	10/1/2024 9:54	SMPL
CCV55	CCV55	23400	5.1673 -		SFAM01.1	PPB	10/1/2024 9:57	SMPL
CCB55	CCB55	797	-0.012 -		SFAM01.1	PPB	10/1/2024 9:59	SMPL