



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/2/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 14:05
In Date: 10/01/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/02/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132681

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4237-01	A4E69	1	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4237-02	A4E70	2	1.00	1.00	2.00	2.00	100.0	P.T. SAMPLE
P4237-04	A4F09	3	1.15	8.68	9.83	5.48	49.9	
P4237-05	A4F10	4	1.15	8.84	9.99	6.27	57.9	
P4237-06	A4F11	5	1.15	8.44	9.59	8.4	85.9	
P4237-07	A4F21	6	1.14	8.60	9.74	6.47	62.0	
P4237-08	A4F22	7	1.15	8.42	9.57	6.34	61.6	
P4237-09	A4F23	8	1.15	8.81	9.96	9.00	89.1	
P4237-10	A4F24	9	1.11	8.77	9.88	5.42	49.1	
P4237-11	A4F26	10	1.15	8.82	9.97	5.97	54.6	
P4237-12	A4F32	11	1.19	8.71	9.9	7.12	68.1	
P4237-14	A4EN6	12	1.18	8.48	9.66	7.16	70.5	
P4237-15	A4EN7	13	1.17	8.51	9.68	7.00	68.5	
P4237-16	A4EN9	14	1.14	8.46	9.6	6.89	68.0	
P4237-17	A4EP6	15	1.12	8.51	9.63	6.04	57.8	
P4237-18	A4EP7	16	1.15	8.40	9.55	7.41	74.5	
P4237-19	A4EQ0	17	1.15	8.81	9.96	5.54	49.8	
P4237-20	A4EQ7	18	1.15	8.80	9.95	6.42	59.9	
P4237-21	VHBLK001	19	1.00	1.00	2.00	2.00	100.0	vhblk

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$