

LB132685


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

43081



A= 0.0000e+000

B= 2.3145e-004 slope

C= 7.6347e-003 y-intercept

Rho= 0.9999622

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	CV1
0.00	0.000	0.016	0.016	38	0.000	38					1
0.05	0.050					442	0				6
0.20	0.200	0.211	0.011	880	0.0 %	880					1
2.50	2.500	2.439	-0.061	10503	0.0 %	10503					2
5.00	5.000	5.022	0.022	21664	0.0 %	21664					0
7.50	7.500	7.533	0.033	32513	0.0 %	32513					0
10.0	10.000	9.979	-0.021	43081	0.0 %	43081					0

LB132685

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	38	-		0 SFAM01.1	PPB	10/1/2024 15:22	Std
	0.2 S01	880	-		0.2 SFAM01.1	PPB	10/1/2024 15:24	Std
	2.5 S02	10503	-		2.5 SFAM01.1	PPB	10/1/2024 15:27	Std
	5 S03	21664	-		5 SFAM01.1	PPB	10/1/2024 15:29	Std
	7.5 S04	32513	-		7.5 SFAM01.1	PPB	10/1/2024 15:31	Std
	10 S05	43081	-		10 SFAM01.1	PPB	10/1/2024 15:33	Std
ICV018	ICV018	16338	3.7891 -		SFAM01.1	PPB	10/1/2024 15:36	SMPL
ICB018	ICB018	-229	-0.0454 -		SFAM01.1	PPB	10/1/2024 15:38	SMPL
CCV056	CCV056	21820	5.058 -		SFAM01.1	PPB	10/1/2024 15:41	SMPL
CCB056	CCB056	-219	-0.0431 -		SFAM01.1	PPB	10/1/2024 15:43	SMPL
PB163829BL	PBW829	-18	0.0035 -		SFAM01.1	PPB	10/1/2024 15:45	SMPL
P4082-01	MC0AP4	2584	0.6057 -		SFAM01.1	PPB	10/1/2024 15:48	SMPL
P4082-02	MC0AP6	-274	-0.0558 -		SFAM01.1	PPB	10/1/2024 15:50	SMPL
P4082-03	MC0AX6	6	0.009 -		SFAM01.1	PPB	10/1/2024 15:52	SMPL
P4082-04	MC0AX6D	59	0.0213 -		SFAM01.1	PPB	10/1/2024 15:54	SMPL
P4082-05	MC0AX6S	4693	1.0938 -		SFAM01.1	PPB	10/1/2024 16:00	SMPL
P4094-01	EIKU0	-83	-0.0116 -		SFAM01.1	PPB	10/1/2024 16:03	SMPL
P4094-02	EIKU1	7	0.0093 -		SFAM01.1	PPB	10/1/2024 16:05	SMPL
P4094-03	EIKU2	87	0.0278 -		SFAM01.1	PPB	10/1/2024 16:07	SMPL
P4094-04	EIKU2D	61	0.0218 -		SFAM01.1	PPB	10/1/2024 16:09	SMPL
P4094-05	EIKU2S	4200	0.9797 -		SFAM01.1	PPB	10/1/2024 16:12	SMPL
P4095-01	E24C3	-97	-0.0148 -		SFAM01.1	PPB	10/1/2024 16:14	SMPL
P4095-02	E24C4	-54	-0.0049 -		SFAM01.1	PPB	10/1/2024 16:16	SMPL
P4095-03	E24C5	-30	0.0007 -		SFAM01.1	PPB	10/1/2024 16:18	SMPL
P4095-04	E24C5D	-30	0.0007 -		SFAM01.1	PPB	10/1/2024 16:21	SMPL
P4095-05	E24C5S	4256	0.9927 -		SFAM01.1	PPB	10/1/2024 16:23	SMPL
CCV057	CCV057	21160	4.9052 -		SFAM01.1	PPB	10/1/2024 16:25	SMPL
CCB057	CCB057	-170	-0.0317 -		SFAM01.1	PPB	10/1/2024 16:27	SMPL