

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132688

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/01/2024	15:25
2	CAL2	1	Water	NA	NA	20.3	7.00	10/01/2024	15:26
3	CAL3	1	Water	NA	NA	20.3	10.02	10/01/2024	15:30
4	ICV	1	Water	NA	NA	20.3	7.02	10/01/2024	15:37
5	CCV1	1	Water	NA	NA	20.3	2.01	10/01/2024	15:40
6	P4235-02	1	Water	NA	NA	20.1	12.07	10/01/2024	15:50
7	P4254-05	1	Water	NA	NA	20.9	6.76	10/01/2024	16:00
8	P4254-05DUP	1	Water	NA	NA	20.8	6.77	10/01/2024	16:01
9	CCV2	1	Water	NA	NA	20.3	12.02	10/01/2024	16:05

WORKLIST(Hardcopy Internal Chain)

VB 132688

WorkList Name : ph p4209-1 WorkList ID : 184001 Department : Wet-Chemistry Date : 10-01-2024 15:09:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4235-02 J	A1403	Water	pH	Cool 4 deg C	PSEG03	J51	09/30/2024	9040C
P4254-05 F	MH-147	Water	pH	Cool 4 deg C	PSEG03	J51	10/01/2024	9040C

Date/Time 10/01/24 15:15
Raw Sample Received by: JD WWC
Raw Sample Relinquished by: JTCsm

Date/Time 10/01/24 17:00
Raw Sample Received by: JTCsm
Raw Sample Relinquished by: JD WWC