

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132690

Slope : 98.6

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/01/2024	16:15
2	CAL2	1	Water	NA	NA	20.3	7.00	10/01/2024	16:20
3	CAL3	1	Water	NA	NA	20.3	10.02	10/01/2024	16:25
4	ICV	1	Water	NA	NA	20.4	7.01	10/01/2024	16:30
5	CCV1	1	Water	NA	NA	20.4	2.01	10/01/2024	16:33
6	P4258-03	1	Solid	20.02	20	20.1	7.64	10/01/2024	16:37
7	P4258-03DUP	1	Solid	20.03	20	20.2	7.65	10/01/2024	16:38
8	CCV2	1	Water	NA	NA	20.3	12.02	10/01/2024	16:40

WORKLIST(Hardcopy Internal Chain)

VB 132690

WorkList Name : ph s p4258 WorkList ID : 184006 Department : Wet-Chemistry Date : 10-01-2024 16:05:16

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4258-03	Chamberlain Ave	Solid	pH	Cool 4 deg C	ROMA02	H11	10/01/2024	9045D

Date/Time 10/01/24 16:10
Raw Sample Received by: JH WXL
Raw Sample Relinquished by: CLS

Date/Time 10/01/24 17:10
Raw Sample Received by: CLS
Raw Sample Relinquished by: JH WXL