



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/3/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:15
In Date: 10/02/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:47
Out Date: 10/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132702

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4243-01	MYD047	1	1.18	8.58	9.76	9.34	95.1	
P4243-02	MYD048	2	1.15	8.35	9.5	9.09	95.1	
P4243-03	MYD056	3	1.18	8.42	9.6	9.34	96.9	
P4243-04	MYD058	4	1.15	8.82	9.97	9.47	94.3	
P4243-05	MYD064	5	1.18	8.53	9.71	9.12	93.1	
P4243-06	MYD065	6	1.19	8.62	9.81	9.26	93.6	
P4243-07	MYD066	7	1.15	8.40	9.55	9.22	96.1	
P4243-08	MYD067	8	1.12	8.48	9.6	9.33	96.8	
P4243-09	MYD068	9	1.15	7.34	8.49	7.76	90.1	
P4243-10	MYD069	10	1.15	8.42	9.57	9.31	96.9	
P4243-11	MYD070	11	1.15	8.57	9.72	9.48	97.2	
P4243-12	MYD071	12	1.18	8.68	9.86	9.41	94.8	
P4243-13	MYD072	13	1.15	8.43	9.58	9.31	96.8	
P4243-14	MYD073	14	1.14	8.62	9.76	9.6	98.1	
P4243-15	MYD074	15	1.19	8.40	9.59	9.17	95.0	
P4243-16	MYD075	16	1.15	8.68	9.83	9.47	95.9	
P4243-17	MYD076	17	1.12	8.65	9.77	9.48	96.6	
P4243-19	MYD078	18	1.12	8.66	9.78	9.34	94.9	
P4243-20	MYD078D	19	1.12	8.66	9.78	9.34	94.9	
P4243-21	MYD078S	20	1.12	8.66	9.78	9.34	94.9	
P4243-22	MYD079	21	1.11	8.77	9.88	9.62	97.0	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$