

LB132017

7470A

INSTRUMENT ID: CV2

Linear

μ Abs.:

40902

A= 0.0000e+000

B= 2.4615e-004

C= -1.1134e-002

Rho= 0.9996602

Accept=Accepted

Max. Conc: 10.000 PPB

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%DD
0.0	0.000	-0.006	-0.006	22	0.000	22	0				-14
0.2	0.200	0.173	-0.027	750	0.0 %	750					1
2.5	2.500	2.518	0.018	10276	0.0 %	10276					3
5.0	5.000	5.137	0.137	20915	0.0 %	20915					-2
7.5	7.500	7.320	-0.180	29783	0.0 %	29783					1
10.0	10.000	10.057	0.057	40902	0.0 %	40902					

LB132717 INSTRUMENT ID : CV1

Method: 7470A

Operator: Admin

Date of Analysis: 02 Oct 2024 17:03:30

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type	Type
0.0 - 1	50	22	-	0.0000	7470A	PPB	02 Oct 2024 17:08:34	S	Std
0.2 - 1	16.2	750	-	0.2000	7470A	PPB	02 Oct 2024 17:15:59	S	Std
2.5 - 1	52.5	10276	-	2.5000	7470A	PPB	02 Oct 2024 17:18:16	S	Std
5.0 - 1	55	20915	-	5.0000	7470A	PPB	02 Oct 2024 17:20:33	S	Std
7.5 - 1	57.5	29783	-	7.5000	7470A	PPB	02 Oct 2024 17:25:26	S	Std
10.0 - 1	511	40902	-	10.0000	7470A	PPB	02 Oct 2024 17:27:43	S	Std
ICV21 - 1	ICV21	15294	3.7535	-	7470A	PPB	02 Oct 2024 17:33:01	U	SMPL
ICB21 - 1	ICB21	-230	-0.0677	-	7470A	PPB	02 Oct 2024 17:35:16	U	SMPL
CCV64 - 1	CCV64	21153	5.1957	-	7470A	PPB	02 Oct 2024 17:37:32	U	SMPL
CCB64 - 1	CCB64	-292	-0.0830	-	7470A	PPB	02 Oct 2024 17:39:46	U	SMPL
CRA - 1	CRA	839	0.1954	-	7470A	PPB	02 Oct 2024 17:44:48	U	SMPL
HighStd - 1	HighStd	40749	10.0192	-	7470A	PPB	02 Oct 2024 17:47:03	U	SMPL
ChkStd - 1	ChkStd	28299	6.9547	-	7470A	PPB	02 Oct 2024 17:49:18	U	SMPL
PB163861BL - 1	PBW	12	-0.0082	-	7470A	PPB	02 Oct 2024 17:55:17	U	SMPL
PB163861BS - 1	LCSW	15869	3.8950	-	7470A	PPB	02 Oct 2024 17:57:32	U	SMPL
P4191-03 - 1	HOPPER-1	341	0.0728	-	7470A	PPB	02 Oct 2024 17:59:48	U	SMPL
P4191-03DUP - 1	HOPPER-1DUP	263	0.0536	-	7470A	PPB	02 Oct 2024 18:04:39	U	SMPL
P4191-03MS - 1	HOPPER-1MS	14479	3.5529	-	7470A	PPB	02 Oct 2024 18:06:55	U	SMPL
P4191-03MSD - 1	HOPPER-1MSD	15546	3.8155	-	7470A	PPB	02 Oct 2024 18:09:13	U	SMPL
P4235-02 - 1	A1403	1137	0.2687	-	7470A	PPB	02 Oct 2024 18:11:32	U	SMPL
CCV65 - 1	CCV65	20790	5.1063	-	7470A	PPB	02 Oct 2024 18:13:48	U	SMPL
CCB65 - 1	CCB65	-215	-0.0641	-	7470A	PPB	02 Oct 2024 18:16:04	U	SMPL
P4191-03LX5 - 1		117	0.0177	-	7470A	PPB	02 Oct 2024 18:20:03	U	SMPL
P4191-03A - 1		14905	3.6577	-	7470A	PPB	02 Oct 2024 18:22:18	U	SMPL
CCV66 - 1	CCV66	21392	5.2545	-	7470A	PPB	02 Oct 2024 18:24:33	U	SMPL
CCB66 - 1	CCB66	-227	-0.0670	-	7470A	PPB	02 Oct 2024 18:26:50	U	SMPL