

LB 132720


SFAM01.1

INSTRUMENT 10: CV1

Linear

 μ Abs.:

43730



A= 0.0000e+000

B= 2.3007e-004 *slope*C= 5.4056e-002 *y-intercept*

Rho= 0.9997525

Accept=Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	%D
0.00	0.000	0.069	0.069	64	0.000	64					=
0.05	0.050					442	0				12
0.20	0.200	0.224	0.024	740	0.0 %	740					-3
2.50	2.500	2.431	-0.069	10332	0.0 %	10332					0
5.00	5.000	4.992	-0.008	21461	0.0 %	21461					-2
7.50	7.500	7.369	-0.131	31796	0.0 %	31796					1
10.0	10.000	10.115	0.115	43730	0.0 %	43730					

LB132720

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	64	-		0 SFAM01.1	PPB	10/3/2024 8:33	Std
	0.2 S01	740	-		0.2 SFAM01.1	PPB	10/3/2024 8:36	Std
	2.5 S02	10332	-		2.5 SFAM01.1	PPB	10/3/2024 8:38	Std
	5 S03	21461	-		5 SFAM01.1	PPB	10/3/2024 8:42	Std
	7.5 S04	31796	-		7.5 SFAM01.1	PPB	10/3/2024 8:44	Std
	10 S05	43730	-		10 SFAM01.1	PPB	10/3/2024 8:46	Std
ICV022	ICV022	16276	3.7986 -		SFAM01.1	PPB	10/3/2024 8:50	SMPL
ICB022	ICB022	-165	0.0161 -		SFAM01.1	PPB	10/3/2024 8:52	SMPL
CCV067	CCV067	22144	5.1487 -		SFAM01.1	PPB	10/3/2024 8:57	SMPL
CCB067	CCB067	-352	-0.0269 -		SFAM01.1	PPB	10/3/2024 8:59	SMPL
PB163862BL	PBW862	148	0.0881 -		SFAM01.1	PPB	10/3/2024 9:04	SMPL
P4108-01	MCOB23	3159	0.7808 -		SFAM01.1	PPB	10/3/2024 9:06	SMPL
P4108-02	MCOB30	121	0.0819 -		SFAM01.1	PPB	10/3/2024 9:08	SMPL
P4108-03	MCOB32	195	0.0989 -		SFAM01.1	PPB	10/3/2024 9:11	SMPL
P4108-04	MCOB33	153	0.0893 -		SFAM01.1	PPB	10/3/2024 9:13	SMPL
P4108-05	MCOB33D	181	0.0957 -		SFAM01.1	PPB	10/3/2024 9:15	SMPL
P4108-06	MCOB33S	4266	1.0355 -		SFAM01.1	PPB	10/3/2024 9:17	SMPL
P4108-07	MCOB42	226	0.1061 -		SFAM01.1	PPB	10/3/2024 9:20	SMPL
P4108-08	MCOB43	62	0.0683 -		SFAM01.1	PPB	10/3/2024 9:22	SMPL
P4108-09	MCOB66	167	0.0925 -		SFAM01.1	PPB	10/3/2024 9:24	SMPL
P4108-10	MCOB68	106	0.0784 -		SFAM01.1	PPB	10/3/2024 9:27	SMPL
P4108-11	MCOB71	58	0.0674 -		SFAM01.1	PPB	10/3/2024 9:29	SMPL
P4108-12	MCOB34	165	0.092 -		SFAM01.1	PPB	10/3/2024 9:31	SMPL
P4108-13	MCOB62	6339	1.5125 -		SFAM01.1	PPB	10/3/2024 9:33	SMPL
P4108-14	MCOB69	-490	-0.0587 -		SFAM01.1	PPB	10/3/2024 9:36	SMPL
P4108-15	MCOB72	247	0.1109 -		SFAM01.1	PPB	10/3/2024 9:38	SMPL
CCV068	CCV068	21069	4.9014 -		SFAM01.1	PPB	10/3/2024 9:40	SMPL
CCB068	CCB068	-73	0.0373 -		SFAM01.1	PPB	10/3/2024 9:42	SMPL