

Analytical Summary Report

Analysis Method: 9040C

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132722

Slope : 98.3

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/03/2024	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	10/03/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/03/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.01	10/03/2024	08:15
5	CCV1	1	Water	NA	NA	20.1	2.02	10/03/2024	09:05
6	P4274-01	1	Water	NA	NA	20.7	7.33	10/03/2024	09:10
7	P4275-02	1	Water	NA	NA	22.1	6.66	10/03/2024	09:15
8	P4278-02	1	Water	NA	NA	20.2	6.85	10/03/2024	08:20
9	P4278-02DUP	1	Water	NA	NA	20.3	6.86	10/03/2024	08:21
10	CCV2	1	Water	NA	NA	20.3	12.02	10/03/2024	08:25

WORKLIST(Hardcopy Internal Chain)

VB 132722

WorkList Name : ph p4278 water WorkList ID : 184061 Department : Wet-Chemistry Date : 10-03-2024 07:53:48

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4274-01	3 222254	Water	pH	Cool 4 deg C	PSEG03	J53	10/02/2024	9040C
P4275-02	C TRE-1768	Water	pH	Cool 4 deg C	PSEG03	J53	10/02/2024	9040C
P4278-02	C BUR-1280	Water	pH	Cool 4 deg C	PSEG03	J53	10/02/2024	9040C

Date/Time 10/03/24 08:00
Raw Sample Received by: RD WLC
Raw Sample Relinquished by: RM SM

Date/Time 10/03/24 11:30
Raw Sample Received by: RM SM
Raw Sample Relinquished by: RD WLC