

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: pH

Supervisor Review By : Iwona

Run Number: LB132723

Slope : 98.3

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.2	4.01	10/03/2024	08:10
2	CAL2	1	Water	NA	NA	20.2	7.00	10/03/2024	08:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/03/2024	08:12
4	ICV	1	Water	NA	NA	20.3	7.01	10/03/2024	08:15
5	CCV1	1	Water	NA	NA	20.2	2.01	10/03/2024	08:20
6	P4273-01	1	Solid	20.02	20	22.3	8.65	10/03/2024	08:25
7	P4275-01	1	Solid	20.03	20	21.7	6.19	10/03/2024	08:35
8	P4277-01	1	Solid	20.02	20	21.9	7.95	10/03/2024	08:40
9	P4278-01	1	Solid	20.03	20	20.4	9.30	10/03/2024	08:49
10	P4278-01DUP	1	Solid	20.04	20	20.5	9.31	10/03/2024	08:50
11	CCV2	1	Water	NA	NA	20.3	12.02	10/03/2024	09:00

WORKLIST(Hardcopy Internal Chain)

VB 132723

WorkList Name : ph p4278 solid WorkList ID : 184062 Department : Wet-Chemistry Date : 10-03-2024 07:54:37

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4273-01	AU-05-100224	Solid	pH	Cool 4 deg C	PSEG03	J53	10/02/2024	9045D
P4275-01	TRE-1300	Solid	pH	Cool 4 deg C	PSEG03	J53	10/02/2024	9045D
P4277-01	BU-03-10022024	Solid	pH	Cool 4 deg C	PSEG05	J53	10/02/2024	9045D
P4278-01	BUR-1300	Solid	pH	Cool 4 deg C	PSEG03	J53	10/02/2024	9045D

Date/Time 10/03/24 08:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]

Date/Time 10/03/24 12:00
Raw Sample Received by: [Signature]
Raw Sample Relinquished by: [Signature]