



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/4/2024

OVENTEMP IN Celsius(°C): 108
Time IN: 12:20
In Date: 10/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:48
Out Date: 10/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132727

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4244-01	MYD080	1	1.17	8.62	9.79	9.55	97.2	
P4244-02	MYD081	2	1.18	8.77	9.95	9.74	97.6	
P4244-03	MYD082	3	1.15	8.55	9.7	9.56	98.4	
P4244-04	MYD083	4	1.16	8.64	9.8	9.4	95.4	
P4244-05	MYD084	5	1.16	8.67	9.83	9.4	95.0	
P4244-06	MYD085	6	1.17	8.56	9.73	9.41	96.3	
P4244-07	MYD086	7	1.14	8.70	9.84	9.14	92.0	
P4244-08	MYD087	8	1.16	8.78	9.94	9.45	94.4	
P4244-09	MYD088	9	1.17	8.71	9.88	9.6	96.8	
P4244-10	MYD089	10	1.16	8.66	9.82	9.28	93.8	
P4244-11	MYD089D	11	1.16	8.66	9.82	9.28	93.8	
P4244-12	MYD089S	12	1.16	8.66	9.82	9.28	93.8	
P4244-13	MYD090	13	1.17	8.79	9.96	9.38	93.4	
P4244-14	MYD091	14	1.15	8.72	9.87	9.6	96.9	
P4244-15	MYD092	15	1.14	8.69	9.83	9.44	95.5	
P4244-16	MYD093	16	1.14	8.70	9.84	9.51	96.2	
P4244-17	MYD094	17	1.15	8.69	9.84	9.65	97.8	
P4244-18	MYD095	18	1.17	8.80	9.97	9.75	97.5	
P4244-19	MYD096	19	1.16	8.71	9.87	9.7	98.0	
P4244-20	MYD097	20	1.15	8.73	9.88	9.75	98.5	
P4244-21	MYD098	21	1.15	8.73	9.88	9.3	93.4	
P4244-22	MYD099	22	1.16	8.56	9.72	9.14	93.2	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$