

PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/4/2024

OVENTEMP IN Celsius(°C): 108
Time IN: 13:00
In Date: 10/03/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 08:00
Out Date: 10/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132728

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4245-01	MYCZS0	1	1.15	8.66	9.81	9.41	95.4	
P4245-02	MYCZS1	2	1.15	8.68	9.83	9.34	94.4	
P4245-03	MYCZS2	3	1.16	8.72	9.88	9.36	94.0	
P4245-04	MYD039	4	1.17	8.73	9.9	9.54	95.9	
P4245-05	MYD040	5	1.14	8.83	9.97	9.51	94.8	
P4245-06	MYD041	6	1.15	7.97	9.12	8.82	96.2	
P4245-07	MYD042	7	1.16	8.65	9.81	9.63	97.9	
P4245-08	MYD043	8	1.14	8.82	9.96	9.51	94.9	
P4245-09	MYD044	9	1.14	8.57	9.71	9.03	92.1	
P4245-10	MYD045	10	1.15	8.72	9.87	9.12	91.4	
P4245-11	MYD0A0	11	1.14	8.82	9.96	9.41	93.8	
P4245-12	MYD0A1	12	1.16	8.69	9.85	9.25	93.1	
P4245-13	MYD0A2	13	1.17	8.71	9.88	9.34	93.8	
P4245-14	MYD0A3	14	1.15	8.77	9.92	9.6	96.4	
P4245-15	MYD0A4	15	1.16	8.67	9.83	9.55	96.8	
P4245-16	MYD0A5	16	1.15	8.65	9.8	9.55	97.1	
P4245-17	MYD0A5D	17	1.15	8.65	9.8	9.55	97.1	
P4245-18	MYD0A5S	18	1.15	8.65	9.8	9.55	97.1	
P4245-19	MYD0A6	19	1.16	8.76	9.92	9.61	96.5	
P4245-20	MYD0A7	20	1.15	8.59	9.74	9.44	96.5	
P4245-21	MYD0A8	21	1.16	8.39	9.55	8.79	90.9	
P4245-22	MYD0A9	22	1.16	8.79	9.95	9.23	91.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$