

LB132740


SFAM01.1

INSTRUMENT ID: CV1

Linear

 μ Abs.:

40477



A= 0.0000e+000

B= 2.5394e-004 slope

C= -2.8420e-001 y-intercept

Rho= 0.9999081

Accept = Accepted

Std ID	Conc.	Calc.	Dev.	Mean	SD or %RSD	Rep 1	Rep 2	Rep 3	Rep 4	Rep 5	G/bD
0.00	0.000	-0.070	-0.070	842	0.000	842					
0.05	0.050					442	0				
0.20	0.200	0.220	0.020	1984	0.0 %	1984					
2.50	2.500	2.593	0.093	11330	0.0 %	11330					
5.00	5.000	4.968	-0.032	20685	0.0 %	20685					
7.50	7.500	7.495	-0.005	30635	0.0 %	30635					
10.0	10.000	9.994	-0.006	40477	0.0 %	40477					

0/00
 1
 10
 4
 -1
 0
 0

LB132740

INSTRUMENT ID : CV1

Sample ID	Extended ID	μ Abs.	Conc.	Std Conc	Method	Units	Date	Type
	0 S0	842	-		0 SFAM01.1	PPB	10/3/2024 15:28	Std
	0.2 S01	1984	-		0.2 SFAM01.1	PPB	10/3/2024 15:31	Std
	2.5 S02	11330	-		2.5 SFAM01.1	PPB	10/3/2024 15:33	Std
	5 S03	20685	-		5 SFAM01.1	PPB	10/3/2024 15:35	Std
	7.5 S04	30635	-		7.5 SFAM01.1	PPB	10/3/2024 15:41	Std
	10 S05	40477	-		10 SFAM01.1	PPB	10/3/2024 15:46	Std
ICV023	ICV023	17515	4.1635 -		SFAM01.1	PPB	10/3/2024 15:48	SMPL
ICB023	ICB023	980	-0.0353 -		SFAM01.1	PPB	10/3/2024 15:51	SMPL
CCV069	CCV069	21323	5.1305 -		SFAM01.1	PPB	10/3/2024 15:53	SMPL
CCB069	CCB069	1119	0 -		SFAM01.1	PPB	10/3/2024 15:55	SMPL
PB163889BL	PBS889	1127	0.002 -		SFAM01.1	PPB	10/3/2024 15:59	SMPL
P4012-01	MCOAT8	2003	0.2244 -		SFAM01.1	PPB	10/3/2024 16:02	SMPL
P4012-02	MCOAT9	460	-0.1674 -		SFAM01.1	PPB	10/3/2024 16:04	SMPL
P4012-03	MCOAW0	962	-0.0399 -		SFAM01.1	PPB	10/3/2024 16:06	SMPL
P4012-04	MCOAW1	1368	0.0632 -		SFAM01.1	PPB	10/3/2024 16:08	SMPL
P4012-05	MCOAW6	9946	2.2414 -		SFAM01.1	PPB	10/3/2024 16:18	SMPL
P4012-06	MCOAW6D	11019	2.5139 -		SFAM01.1	PPB	10/3/2024 16:21	SMPL
P4012-07	MCOAW6S	21834	5.2602 -		SFAM01.1	PPB	10/3/2024 16:24	SMPL
P4012-08	MCOAW9	6652	1.405 -		SFAM01.1	PPB	10/3/2024 16:26	SMPL
P4012-09	MCOAX0	1901	0.1985 -		SFAM01.1	PPB	10/3/2024 16:28	SMPL
CCV070	CCV070	21661	5.2163 -		SFAM01.1	PPB	10/3/2024 16:30	SMPL
CCB070	CCB070	1229	0.0279 -		SFAM01.1	PPB	10/3/2024 16:36	SMPL