

Analytical Summary Report

Analysis Method: 9045D

Analyst By : jignesh

Parameter: Corrosivity

Supervisor Review By : Iwona

Run Number: LB132746

Slope : 98.4

BalanceID: WC SC-4

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/03/2024	15:10
2	CAL2	1	Water	NA	NA	20.3	7.00	10/03/2024	15:11
3	CAL3	1	Water	NA	NA	20.3	10.02	10/03/2024	15:15
4	ICV	1	Water	NA	NA	20.3	7.00	10/03/2024	15:20
5	CCV1	1	Water	NA	NA	20.3	2.01	10/03/2024	15:25
6	P4270-04	1	Solid	20.02	20	21.7	8.47	10/03/2024	15:35
7	P4270-04DUP	1	Solid	20.03	20	21.8	8.48	10/03/2024	15:36
8	P4276-02	1	Solid	20.02	20	22.1	8.81	10/03/2024	15:45
9	P4285-01	1	Solid	20.03	20	22.4	7.81	10/03/2024	15:50
10	P4290-04	1	Solid	20.02	20	22.6	8.80	10/03/2024	16:00
11	CCV2	1	Water	NA	NA	20.3	12.02	10/03/2024	16:10

WORKLIST(Hardcopy Internal Chain)

VB 132746

WorkList Name : corrosivity p4290 WorkList ID : 184099 Department : Wet-Chemistry Date : 10-03-2024 14:30:52

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4270-04	IB-2	Solid	Corrosivity	Cool 4 deg C	PSEG03	J51	10/02/2024	9045D
P4276-02	VNJ-208	Solid	Corrosivity	Cool 4 deg C	PSEG03	J53	10/02/2024	9045D
P4285-01	OILY-DEBRIS-DRUM	Solid	Corrosivity	Cool 4 deg C	PSEG03	J51	10/03/2024	9045D
P4290-04	IB-4	Solid	Corrosivity	Cool 4 deg C	PSEG03	J53	10/03/2024	9045D

Date/Time 10/03/24 15:00

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

Date/Time 10/03/24

Raw Sample Received by: [Signature]

Raw Sample Relinquished by: [Signature]

17130

[Signature]

[Signature]