

Analytical Summary Report

Analysis Method: 1010B

Reviewed By: rubina

Parameter: Flash Point

Supervisor Review By: Iwona

Run Number: LB132749

Ambient Barometric Pressure (mmHg): 771.00

Thermometer ID: Flash Point

Barometric Scale ID: 0511064

Reagent/Standard	Lot/Log #
p-xylene (ICV)	W3088

Seq	LabID	True Value °F	DL	Initial Sample °C	Celsius °C	Result °F	Final Result °F	Anal Date	Anal Time
1	ICV	81	1	8	28.00	82.4	81.7	10/03/2024	08:30
2	P4254-05		1	12	100.00	>212.0	>212.0	10/03/2024	09:00
3	P4254-05DUP		1	12	100.00	>212.0	>212.0	10/03/2024	09:30
4	P4274-01		1	14	100.00	>212.0	>212.0	10/03/2024	10:00
5	P4275-02		1	11	100.00	>212.0	>212.0	10/03/2024	10:30
6	P4278-02		1	14	100.00	>212.0	>212.0	10/03/2024	11:00

Result = (Celsius * 1.8) + 32

Final Result = Result + (760 - Ambient Barometric Pressure) * 0.06

6132749

WORKLIST(Hardcopy Internal Chain)

WorkList Name : fp-10-3

WorkList ID : 184080

Department : Wet-Chemistry

Date : 10-03-2024 08:11:23

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4254-05 M	MH-147	Water	Flash Point	Cool 4 deg C	PSEG03	J51	10/01/2024	1010B
P4274-01 R	222254	Water	Flash Point	Cool 4 deg C	PSEG03	J53	10/02/2024	1010B
P4275-02 E	TRE-1768	Water	Flash Point	Cool 4 deg C	PSEG03	J53	10/02/2024	1010B
P4278-02 E	BUR-1280	Water	Flash Point	Cool 4 deg C	PSEG03	J53	10/02/2024	1010B

Date/Time 10/03/2024 08:20
Raw Sample Received by: RMCW
Raw Sample Relinquished by: RMCW

Date/Time 10/03/2024 11:40
Raw Sample Received by: RMCW
Raw Sample Relinquished by: RMCW