

Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: jignesh

Run Number: LB132750

Reagent/Standard	Lot/Log #
Residual chlorine ICV-LCS, 0.4PPM	WP110044
Chlorine Calibration std, 0.1ppm	WP110039
Chlorine Calibration std, 0.2ppm	WP110040
Chlorine Calibration std, 0.8ppm	WP110041
Chlorine Calibration std, 0.0ppm	WP110038
Chlorine Calibration std, 1.6ppm	WP110042
Residual Chlorine Calibration and CCV std, 0	WP110043
Total Chlorine Powder Pillows	W3147

Intercept: 0.0023

Slope: 0.9924

Regression: 0.999726

Seq	Lab ID	True Val (mg/l)	DF	Initial Reading	Final Reading	Difference	Result (mg/l)	%D	AnalDate	Anal Time
1	CAL1	0	1	0.000	0.000	0.000	0.00		10/03/2024	17:00
2	CAL2	0.1	1	0.000	0.110	0.110	0.11	9	10/03/2024	17:03
3	CAL3	0.2	1	0.000	0.200	0.200	0.20	-0.5	10/03/2024	17:06
4	CAL4	0.4	1	0.000	0.410	0.410	0.41	2.7	10/03/2024	17:09
5	CAL5	0.8	1	0.000	0.770	0.770	0.77	-3.3	10/03/2024	17:12
6	CAL6	1.6	1	0.000	1.600	1.600	1.61	0.6	10/03/2024	17:15



Analytical Summary Report

Analysis Method: SM4500 Cl G

ANALYST: Iwona

Parameter: Residual Chlorine

SUPERVISOR REVIEW BY: jignesh

Run Number: LB132750

Seq	Lab ID	Initial Weight	Final Vol	True Value (mg/L)	DF	Initial Reading	Final Reading	Diff.	Result (mg/L)	Anal Date	Anal Time
1	ICV			0.4	1	0.0000	0.4000	0.4000	0.4010	10/03/2024	17:18
2	ICB				1	0.0000	0.0100	0.0100	0.0080	10/03/2024	17:21
3	CCV1			0.4	1	0.0000	0.4100	0.4100	0.4110	10/03/2024	17:24
4	CCB1				1	0.0000	0.0000	0.0000	-0.0020	10/03/2024	17:27
5	LB132750BL	50	50		1	0.0000	0.0000	0.0000	-0.0020	10/03/2024	17:30
6	LB132750BS	50	50	0.4	1	0.0000	0.4000	0.4000	0.4010	10/03/2024	17:33
7	P4286-01	50	50		1	0.0000	0.5400	0.5400	0.5420	10/03/2024	17:36
8	P4286-01DUP	50	50		1	0.0000	0.5400	0.5400	0.5420	10/03/2024	17:39
9	P4286-01MS	50	50	0.4	1	0.0000	0.8900	0.8900	0.8940	10/03/2024	17:42
10	P4286-01MSD	50	50	0.4	1	0.0000	0.8800	0.8800	0.8840	10/03/2024	17:45
11	CCV2			0.4	1	0.0000	0.4100	0.4100	0.4110	10/03/2024	17:48
12	CCB2				1	0.0000	0.0000	0.0000	-0.0020	10/03/2024	17:51

LB 132750

WORKLIST(Hardcopy Internal Chain)

WorkList Name : RESCHLORINE-100324

WorkList ID : 184102

Department : Wet-Chemistry

Date : 10-03-2024 14:49:36

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4286-01	Water Treatment Discharge	Water	Residual Chlorine	Cool 4 deg C	VERI01	H11	10/03/2024	SM4500 Cl G

Date/Time 10/03/2024 16:40
Raw Sample Received by: 12/2017
Raw Sample Relinquished by: 25/1017

Date/Time 10/03/2024 18:00
Raw Sample Received by: 28/1017
Raw Sample Relinquished by: 12/2017