

Analytical Summary Report

Analysis Method: 9040C

Parameter: pH

Run Number: LB132757

Analyst By : jignesh

Supervisor Review By : Iwona

Slope : 98.6

pH Meter ID : WC PH METER-1

Calibration Standards	Chemtech Log#
PH 4 BUFFER SOLUTION	W3107
BUFFER PH 7.00 GREEN 1PINT PK6	W3093
PH 10.01 BUFFER,COLOR CD 475ML	W3094
buffer solution pH 7 yellow	W3071
Buffer Solution, PH2 (500ml)	W3005
Buffer Solution, PH12 (500ml)	W3072

True Value of ICV = 7.00 Control Limits[+/- 0.1].

True Value of CCV1 = 2.00 Control Limits[+/- 0.1].

True Value of CCV2 = 12.00 Control Limits[+/- 0.1].

Seq	LabID	DF	Matrix	Weight (gm)	Volume (ml)	Temperature (°C)	Result (pH)	Anal Date	Anal Time
1	CAL1	1	Water	NA	NA	20.3	4.01	10/04/2024	09:44
2	CAL2	1	Water	NA	NA	20.3	7.00	10/04/2024	09:46
3	CAL3	1	Water	NA	NA	20.2	10.02	10/04/2024	09:48
4	ICV	1	Water	NA	NA	20.3	7.00	10/04/2024	09:50
5	CCV1	1	Water	NA	NA	20.3	2.01	10/04/2024	09:55
6	P4285-03	1	Water	NA	NA	22.4	7.81	10/04/2024	10:00
7	P4285-03DUP	1	Water	NA	NA	22.5	7.82	10/04/2024	10:01
8	CCV2	1	Water	NA	NA	20.3	12.02	10/04/2024	10:05

WORKLIST(Hardcopy Internal Chain)

132757

WorkList Name : PH WATER P4285 WorkList ID : 184120 Department : Wet-Chemistry Date : 10-04-2024 09:29:43

Sample	Customer Sample	Matrix	Test	Preservative	Customer	Raw Sample Storage Location	Collect Date	Method
P4285-03	WATER-TOTES	Water	pH	Cool 4 deg C	PSEG03	J51	10/03/2024	9040C

Date/Time 10/04/24 09:35
Raw Sample Received by: JH well
Raw Sample Relinquished by: RM SM

10
Date/Time 10/09/24 10:42
Raw Sample Received by: JH SM
Raw Sample Relinquished by: JH well