



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/7/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 12:35
In Date: 10/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:45
Out Date: 10/05/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132758

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4247-01	MYD9N1	1	1.14	8.36	9.5	9.17	96.1	
P4247-02	MYD9N2	2	1.14	8.80	9.94	9.82	98.6	
P4247-03	MYD9N2D	3	1.14	8.80	9.94	9.82	98.6	
P4247-04	MYD9N2S	4	1.14	8.80	9.94	9.82	98.6	
P4247-05	MYD9N3	5	1.18	8.62	9.8	9.57	97.3	
P4247-06	MYD9N4	6	1.15	8.58	9.73	9.32	95.2	
P4247-07	MYD9N5	7	1.15	8.55	9.7	9.44	97.0	
P4247-08	MYD9N6	8	1.18	8.44	9.62	9.27	95.9	
P4247-09	MYD9N7	9	1.12	8.65	9.77	9.45	96.3	
P4247-10	MYD9N8	10	1.12	8.66	9.78	9.6	97.9	
P4247-11	MYD9N9	11	1.12	8.77	9.89	9.57	96.4	
P4247-12	MYD9P0	12	1.14	8.61	9.75	9.36	95.5	
P4247-13	MYD9P1	13	1.19	8.52	9.71	9.39	96.2	
P4247-14	MYD9P2	14	1.12	8.69	9.81	9.55	97.0	
P4247-15	MYD9P3	15	1.15	8.42	9.57	9.12	94.7	
P4247-16	MYD9P4	16	1.14	8.83	9.97	8.33	81.4	
P4247-17	MYD9P5	17	1.15	8.40	9.55	9.08	94.4	
P4247-18	MYD9P6	18	1.18	8.42	9.6	9.11	94.2	
P4247-19	MYD9P7	19	1.15	8.42	9.57	8.97	92.9	
P4247-20	MYD9P8	20	1.18	8.49	9.67	9.36	96.3	
P4247-21	MYD9P9	21	1.15	8.45	9.6	9.42	97.9	
P4247-22	MYD9Q0	22	1.12	8.82	9.94	9.66	96.8	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$