



PERCENT SOLID

Supervisor: Iwona
Analyst: jignesh
Date: 10/7/2024

OVENTEMP IN Celsius(°C): 107
Time IN: 13:20
In Date: 10/04/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
OvenID: M OVEN#1

OVENTEMP OUT Celsius(°C): 103
Time OUT: 07:59
Out Date: 10/05/2024
Weight Check 1.0g: 1.00
Weight Check 10g: 10.00
BalanceID: M SC-4
Thermometer ID: % SOLID- OVEN

QC:LB132759

Lab ID	Client SampleID	Dish #	Dish Wt(g) (A)	Sample Wt(g)	Dish + Sample Wt(g) (B)	Dish+Dry Sample Wt(g) (C)	% Solid	Comments
P4248-01	MYD9Q1	1	1.14	8.60	9.74	9.57	98.0	
P4248-02	MYD9Q2	2	1.18	8.43	9.61	9.36	97.0	
P4248-03	MYD9Q3	3	1.18	8.52	9.7	9.47	97.3	
P4248-04	MYD9Q4	4	1.17	8.43	9.6	9.36	97.2	
P4248-05	MYD9Q5	5	1.18	8.72	9.9	9.66	97.2	
P4248-06	MYD9Q6	6	1.13	8.74	9.87	9.72	98.3	
P4248-07	MYD9Q7	7	1.16	8.80	9.96	9.84	98.6	
P4248-08	MYD9Q8	8	1.18	8.56	9.74	9.56	97.9	
P4248-09	MYD9Q9	9	1.18	8.62	9.8	9.54	97.0	
P4248-10	MYD9R0	10	1.19	8.41	9.6	9.44	98.1	
P4248-11	MYD9R1	11	1.12	8.71	9.83	9.71	98.6	
P4248-12	MYD9R2	12	1.18	8.46	9.64	9.45	97.8	
P4248-13	MYD9R3	13	1.18	8.74	9.92	9.75	98.1	
P4248-14	MYD9R4	14	1.18	8.45	9.63	9.57	99.3	
P4248-15	MYD9R5	15	1.18	8.74	9.92	9.75	98.1	
P4248-16	MYD9R6	16	1.18	8.72	9.9	9.77	98.5	
P4248-17	MYD9R7	17	1.15	8.75	9.9	9.78	98.6	
P4248-18	MYD9R8	18	1.16	8.61	9.77	9.47	96.5	
P4248-19	MYD9R9	19	1.18	8.72	9.9	9.77	98.5	
P4248-20	MYD9S0	20	1.12	8.47	9.59	9.51	99.1	
P4248-21	MYD9S0D	21	1.12	8.47	9.59	9.51	99.1	
P4248-22	MYD9S0S	22	1.12	8.47	9.59	9.51	99.1	

$$\% \text{ Solid} = \frac{(C-A) * 100}{(B-A)}$$